

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061720\
 Data File : PS010831.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Jun 2020 21:11
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 18 01:34:03 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052920.M
 Quant Title : 8080.M
 QLast Update : Fri May 29 15:43:50 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.228	6.920	661.9E6	432.0E6	836.920	791.975
Target Compounds							
1) T	Dalapon	3.041	2.627	723.4E6	488.0E6	735.733	666.499
2) T	3,5-DICHL...	6.441	6.070	849.9E6	601.0E6	777.465	733.261
3) T	4-Nitroph...	7.058	6.525	288.2E6	241.4E6	777.303	728.820
5) T	DICAMBA	7.420	7.089	2486.0E6	1662.7E6	790.582	738.696
6) T	MCPD	7.590	7.187	193.6E6	139.7E6	78.377	74.993
7) T	MCPA	7.740	7.403	271.8E6	204.4E6	77.740	75.202
8) T	DICHLORPROP	8.112	7.715	595.5E6	430.5E6	785.643	746.663
9) T	2,4-D	8.340	8.008	626.0E6	532.5E6	794.179	702.296
10) T	Pentachlo...	8.653	8.439	8550.2E6	5054.2E6	801.622	711.409
11) T	2,4,5-TP ...	9.201	8.800	3414.1E6	2257.4E6	798.278	751.633
12) T	2,4,5-T	9.486	9.179	3273.6E6	2268.6E6	804.904	755.966
13) T	2,4-DB	10.039	9.684	438.7E6	313.3E6	793.346	771.256
14) T	DINOSEB	11.201	10.019	2521.2E6	1667.3E6	780.293	774.979
15) T	Picloram	11.019	11.000	3682.2E6	2802.4E6	721.282	679.221
16) T	DCPA	11.466	10.948	4179.8E6	2830.6E6	780.102	763.328

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061720\
Data File : PS010831.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Jun 2020 21:11
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 18 01:34:03 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052920.M
Quant Title : 8080.M
QLast Update : Fri May 29 15:43:50 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

