

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061721\
 Data File : PS015714.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Jun 2021 15:48
 Operator : DD\AJ
 Sample : M2672-26
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 18 01:09:34 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061521.M
 Quant Title : 8080.M
 QLast Update : Tue Jun 15 18:15:49 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.191	6.769	936.1E6	290.4E6	563.444	576.278
Target Compounds							
1) T	Dalapon	2.976f	2.484	30746257	45847559	15.233	70.591 #
2) T	3,5-DICHL...	0.000	5.984	0	88819202	N.D.	117.772
3) T	4-Nitroph...	7.029	0.000	924561	0	1.017	N.D. #
5) T	DICAMBA	7.369	0.000	14480899	0	2.252	N.D. #
6) T	MCP	7.548	0.000	3052939	0	<MDL	N.D. #
8) T	DICHLORPROP	8.019f	7.571	9987451	21478101	6.308	41.484 #
9) T	2,4-D	8.371f	0.000	54836405	0	31.289	N.D. #
10) T	Pentachlo...	8.679f	0.000	10765783	0	<MDL	N.D. #
12) T	2,4,5-T	9.398f	9.081f	44827089	10966425	5.078	3.888
13) T	2,4-DB	9.968	0.000	53725444	0	26.265	N.D. #
14) T	DINOSEB	11.122f	9.886	61714305	149.8E6	9.529	74.353 #
15) T	Picloram	10.954	0.000	184.8E6	0	13.514	N.D. #
16) T	DCPA	11.492f	0.000	10668575	0	<MDL	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061721\
Data File : PS015714.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Jun 2021 15:48
Operator : DD\AJ
Sample : M2672-26
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 18 01:09:34 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061521.M
Quant Title : 8080.M
QLast Update : Tue Jun 15 18:15:49 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 2.976 min
Delta R.T.: -0.044 min
Response: 30746257
Conc: 15.23 ng/ml

Instrument :
ECD_S
ClientSampleId :




#1 Dalapon

R.T.: 2.484 min
Delta R.T.: -0.037 min
Response: 45847559
Conc: 70.59 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 0.000 min
Exp R.T. : 6.410 min
Response: 0
Conc: N.D.



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 5.984 min
Delta R.T.: 0.024 min
Response: 88819202
Conc: 117.77 ng/ml



#3 4-Nitrophenol


R.T.: 7.029 min
Delta R.T.: 0.010 min
Response: 924561
Conc: 1.02 ng/ml

Instrument :
ECD_S
ClientSampleId :




#3 4-Nitrophenol

R.T.: 0.000 min
Exp R.T. : 6.389 min
Response: 0
Conc: N.D.




#4 2,4-DCAA

R.T.: 7.191 min
Delta R.T.: -0.002 min
Response: 936131747
Conc: 563.44 ng/ml



#4 2,4-DCAA


R.T.: 6.769 min
Delta R.T.: -0.002 min
Response: 290358833
Conc: 576.28 ng/ml



#5 DICAMBA


R.T.: 7.369 min
Delta R.T.: -0.016 min
Response: 14480899
Conc: 2.25 ng/ml

Instrument :
ECD_S
ClientSampleId :




#5 DICAMBA

R.T.: 0.000 min
Exp R.T. : 6.935 min
Response: 0
Conc: N.D.




#8 DICHLORPROP

R.T.: 8.019 min
Delta R.T.: -0.060 min
Response: 9987451
Conc: 6.31 ng/ml



#8 DICHLORPROP


R.T.: 7.571 min
Delta R.T.: 0.016 min
Response: 21478101
Conc: 41.48 ng/ml



#9 2,4-D


R.T.: 8.371 min
Delta R.T.: 0.066 min
Response: 54836405
Conc: 31.29 ng/ml

Instrument :
ECD_S
ClientSampleId :




#9 2,4-D

R.T.: 0.000 min
Exp R.T. : 7.845 min
Response: 0
Conc: N.D.




#12 2,4,5-T

R.T.: 9.398 min
Delta R.T.: -0.052 min
Response: 44827089
Conc: 5.08 ng/ml



#12 2,4,5-T


R.T.: 9.081 min
Delta R.T.: 0.068 min
Response: 10966425
Conc: 3.89 ng/ml



#13 2,4-DB


R.T.: 9.968 min
Delta R.T.: -0.032 min
Response: 53725444
Conc: 26.26 ng/ml

Instrument :
ECD_S
ClientSampleId :




#13 2,4-DB

R.T.: 0.000 min
Exp R.T. : 9.519 min
Response: 0
Conc: N.D.




#14 DINOSEB

R.T.: 11.122 min
Delta R.T.: -0.041 min
Response: 61714305
Conc: 9.53 ng/ml



#14 DINOSEB


R.T.: 9.886 min
Delta R.T.: 0.039 min
Response: 149757022
Conc: 74.35 ng/ml



#15 Picloram

R.T.: 10.954 min
Delta R.T.: -0.024 min
Response: 184752278
Conc: 13.51 ng/ml

Instrument :
ECD_S
ClientSampleId :



#15 Picloram

R.T.: 0.000 min
Exp R.T. : 10.824 min
Response: 0
Conc: N.D.