

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061721\
 Data File : PS015732.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Jun 2021 23:03
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 111 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 18 01:13:20 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061521.M
 Quant Title : 8080.M
 QLast Update : Tue Jun 15 18:15:49 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.191	6.769	1287.4E6	393.9E6	774.882	781.735
Target Compounds							
1) T	Dalapon	3.020	2.521	1495.7E6	443.4E6	741.043	682.678
2) T	3,5-DICHL...	6.408	5.959	1677.8E6	550.8E6	747.529	730.290
3) T	4-Nitroph...	7.018	6.388	659.7E6	217.1E6	725.615	709.559
5) T	DICAMBA	7.383	6.933	4594.3E6	1549.6E6	714.625	746.799
6) T	MCP	7.553	7.030	364.5E6	123.7E6	67.111	74.207
7) T	MCPA	7.702	7.242	510.7E6	182.9E6	65.239	71.304
8) T	DICHLORPROP	8.076	7.553	1104.6E6	379.2E6	697.603	732.502
9) T	2,4-D	8.303	7.843	1208.4E6	453.6E6	689.528	731.232
10) T	Pentachlo...	8.614	8.272	15310.4E6	5190.5E6	709.363	738.999
11) T	2,4,5-TP ...	9.164	8.633	5813.0E6	2062.8E6	665.010	747.212
12) T	2,4,5-T	9.447	9.010	5838.8E6	2101.6E6	661.367	745.131
13) T	2,4-DB	9.996	9.515	1253.8E6	277.3E6	612.933	742.359
14) T	DINOSEB	11.158	9.844	3983.6E6	1479.3E6	615.094	734.439
15) T	Picloram	10.973	10.821	7381.2E6	2864.0E6	539.919	681.004 #
16) T	DCPA	11.425	10.775	7956.0E6	3158.1E6	601.521	751.179

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061721\
Data File : PS015732.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Jun 2021 23:03
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 111 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 18 01:13:20 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061521.M
Quant Title : 8080.M
QLast Update : Tue Jun 15 18:15:49 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

