

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061722\
 Data File : PS019739.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Jun 2022 09:31
 Operator : AJ\MA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 17 12:31:20 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052422.M
 Quant Title : 8080.M
 QLast Update : Wed May 25 02:05:34 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.197	6.780	1828.1E6	611.2E6	761.959	788.766
Target Compounds							
1) T	Dalapon	3.014	2.526	3370.6E6	725.5E6	664.974	723.014
2) T	3,5-DICHL...	6.412	5.961	2250.7E6	896.2E6	712.098	747.655
3) T	4-Nitroph...	7.030	6.400	774.4E6	313.2E6	789.185	731.800
5) T	DICAMBA	7.386	6.945	6813.4E6	2601.4E6	717.367	760.980
6) T	MCPD	7.556	7.042	545.7E6	204.5E6	72.704	75.533
7) T	MCPA	7.704	7.254	765.4E6	284.8E6	70.847	74.104
8) T	DICHLORPROP	8.077	7.564	1788.9E6	661.4E6	698.610	748.673
9) T	2,4-D	8.307	7.856	2052.5E6	797.1E6	717.769	747.299
10) T	Pentachlo...	8.613	8.281	24432.6E6	8773.8E6	706.371	752.709
11) T	2,4,5-TP ...	9.163	8.642	11025.0E6	3817.0E6	702.186	759.280
12) T	2,4,5-T	9.449	9.019	13113.4E6	3565.6E6	696.795	758.114
13) T	2,4-DB	10.001	9.524	2996.7E6	466.4E6	740.661	754.797
14) T	DINOSEB	11.154	9.848	11972.3E6	2682.3E6	774.047	755.024
15) T	Picloram	10.980	10.831	23028.6E6	4724.1E6	775.968	734.432
16) T	DCPA	11.422	10.777	18716.1E6	4663.0E6	776.471	755.447

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061722\
Data File : PS019739.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Jun 2022 09:31
Operator : AJ\MA
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 17 12:31:20 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052422.M
Quant Title : 8080.M
QLast Update : Wed May 25 02:05:34 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

