

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061819\
 Data File : PS005111.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Jun 2019 01:55
 Operator : SM\AJ
 Sample : K3421-04
 Misc :
 ALS Vial : 67 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 SAMPLE-1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 19 02:17:29 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061219.M
 Quant Title : 8080.M
 QLast Update : Fri Jun 14 09:57:49 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.600	5.314	313.2E6	233.1E6	237.635	233.649
Target Compounds							
1) T	Dalapon	2.096	0.000	27480116	0	15.776	N.D. #
2) T	3,5-DICHL...	0.000	4.632	0	122.2E6	N.D.	95.446
5) T	DICAMBA	5.700	0.000	14626245	0	2.584	N.D. #
6) T	MCP	5.825	0.000	23009234	0	5.158	N.D. #
10) T	Pentachlo...	0.000	6.580	0	29730718	N.D.	2.308
11) T	2,4,5-TP ...	0.000	6.992	0	125.1E6	N.D.	23.365
12) T	2,4,5-T	7.618	0.000	74891525	0	9.373	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061819\
Data File : PS005111.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Jun 2019 01:55
Operator : SM\AJ
Sample : K3421-04
Misc :
ALS Vial : 67 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
SAMPLE-1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 19 02:17:29 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061219.M
Quant Title : 8080.M
QLast Update : Fri Jun 14 09:57:49 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 2.096 min
Delta R.T.: 0.000 min
Response: 27480116
Conc: 15.78 ng/ml

Instrument :
ECD_S
ClientSampleId :
SAMPLE-1




#1 Dalapon

R.T.: 0.000 min
Exp R.T. : 1.794 min
Response: 0
Conc: N.D.




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 0.000 min
Exp R.T. : 4.922 min
Response: 0
Conc: N.D.



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 4.632 min
Delta R.T.: 0.054 min
Response: 122150867
Conc: 95.45 ng/ml



#4 2,4-DCAA


R.T.: 5.600 min
Delta R.T.: 0.026 min
Response: 313154044
Conc: 237.63 ng/ml

Instrument :
ECD_S
ClientSampleId :
SAMPLE-1




#4 2,4-DCAA

R.T.: 5.314 min
Delta R.T.: 0.029 min
Response: 233072916
Conc: 233.65 ng/ml




#5 DICAMBA

R.T.: 5.700 min
Delta R.T.: -0.031 min
Response: 14626245
Conc: 2.58 ng/ml



#5 DICAMBA


R.T.: 0.000 min
Exp R.T.: 5.417 min
Response: 0
Conc: N.D.



#6 MCP


R.T.: 5.825 min
Delta R.T.: -0.060 min
Response: 23009234
Conc: 5.16 ug/ml

Instrument :
ECD_S
ClientSampleId :
SAMPLE-1




#6 MCP

R.T.: 0.000 min
Exp R.T. : 5.500 min
Response: 0
Conc: N.D.




#10 Pentachlorophenol

R.T.: 0.000 min
Exp R.T. : 6.856 min
Response: 0
Conc: N.D.



#10 Pentachlorophenol


R.T.: 6.580 min
Delta R.T.: -0.042 min
Response: 29730718
Conc: 2.31 ng/ml



#11 2,4,5-TP (SILVEX)


R.T.: 0.000 min
Exp R.T.: 7.395 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :
SAMPLE-1




#11 2,4,5-TP (SILVEX)

R.T.: 6.992 min
Delta R.T.: 0.007 min
Response: 125064589
Conc: 23.36 ng/ml



#12 2,4,5-T

R.T.: 7.618 min
Delta R.T.: -0.056 min
Response: 74891525
Conc: 9.37 ng/ml



#12 2,4,5-T

R.T.: 0.000 min
Exp R.T.: 7.349 min
Response: 0
Conc: N.D.