

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061820\
 Data File : PS010839.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Jun 2020 20:56
 Operator : DD\AJ
 Sample : L2811-04
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 NB-08-061620

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 19 02:48:05 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052920.M
 Quant Title : 8080.M
 QLast Update : Fri May 29 15:43:50 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.228	6.920	284.6E6	208.8E6	359.853	382.764
Target Compounds							
1) T	Dalapon	2.994	2.583	36496844	135.2E6	37.121	184.707 #
2) T	3,5-DICHL...	0.000	6.100	0	75851615	N.D.	92.542
3) T	4-Nitroph...	7.115	6.577	18123543	5744228	48.886	17.340 #
5) T	DICAMBA	7.492	7.054	7114250	7316534	2.262	3.251 #
7) T	MCPA	7.688	7.457	87455517	80266456	25.016	29.534
8) T	DICHLORPROP	0.000	7.738	0	11248573	N.D.	19.507
12) T	2,4,5-T	9.433	0.000	39369442	0	9.680	N.D. #
14) T	DINOSEB	11.167	0.000	39724710	0	12.295	N.D. #
15) T	Picloram	10.993	0.000	103.3E6	0	20.241	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061820\
Data File : PS010839.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Jun 2020 20:56
Operator : DD\AJ
Sample : L2811-04
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
NB-08-061620

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 19 02:48:05 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052920.M
Quant Title : 8080.M
QLast Update : Fri May 29 15:43:50 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 2.994 min
Delta R.T.: -0.047 min
Response: 36496844
Conc: 37.12 ng/ml

Instrument :
ECD_S
ClientSampleId :
NB-08-061620




#1 Dalapon

R.T.: 2.583 min
Delta R.T.: -0.041 min
Response: 135231889
Conc: 184.71 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 0.000 min
Exp R.T. : 6.443 min
Response: 0
Conc: N.D.



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 6.100 min
Delta R.T.: 0.030 min
Response: 75851615
Conc: 92.54 ng/ml



#3 4-Nitrophenol


R.T.: 7.115 min
Delta R.T.: 0.055 min
Response: 18123543
Conc: 48.89 ng/ml

Instrument :
ECD_S
ClientSampleId :
NB-08-061620




#3 4-Nitrophenol

R.T.: 6.577 min
Delta R.T.: 0.052 min
Response: 5744228
Conc: 17.34 ng/ml




#4 2,4-DCAA

R.T.: 7.228 min
Delta R.T.: -0.003 min
Response: 284594223
Conc: 359.85 ng/ml



#4 2,4-DCAA


R.T.: 6.920 min
Delta R.T.: -0.002 min
Response: 208765571
Conc: 382.76 ng/ml



#5 DICAMBA


R.T.: 7.492 min
Delta R.T.: 0.069 min
Response: 7114250
Conc: 2.26 ng/ml

Instrument :
ECD_S
ClientSampleId :
NB-08-061620




#5 DICAMBA

R.T.: 7.054 min
Delta R.T.: -0.037 min
Response: 7316534
Conc: 3.25 ng/ml




#7 MCPA

R.T.: 7.688 min
Delta R.T.: -0.063 min
Response: 87455517
Conc: 25.02 ug/ml



#7 MCPA


R.T.: 7.457 min
Delta R.T.: 0.043 min
Response: 80266456
Conc: 29.53 ug/ml



#8 DICHLORPROP


R.T.: 0.000 min
Exp R.T.: 8.116 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :
NB-08-061620




#8 DICHLORPROP

R.T.: 7.738 min
Delta R.T.: 0.021 min
Response: 11248573
Conc: 19.51 ng/ml




#11 2,4,5-TP (SILVEX)

R.T.: 9.267 min
Delta R.T.: 0.060 min
Response: 9422303
Conc: 2.20 ng/ml



#11 2,4,5-TP (SILVEX)


R.T.: 8.843 min
Delta R.T.: 0.041 min
Response: -22706800
Conc: N.D.



#12 2,4,5-T


R.T.: 9.433 min
Delta R.T.: -0.059 min
Response: 39369442
Conc: 9.68 ng/ml

Instrument :
ECD_S
ClientSampleId :
NB-08-061620




#12 2,4,5-T

R.T.: 0.000 min
Exp R.T. : 9.181 min
Response: 0
Conc: N.D.




#14 DINOSEB

R.T.: 11.167 min
Delta R.T.: -0.040 min
Response: 39724710
Conc: 12.29 ng/ml



#14 DINOSEB


R.T.: 0.000 min
Exp R.T. : 10.022 min
Response: 0
Conc: N.D.



#15 Picloram

R.T.: 10.993 min
Delta R.T.: -0.031 min
Response: 103334809
Conc: 20.24 ng/ml

Instrument :
ECD_S
ClientSampleId :
NB-08-061620



#15 Picloram

R.T.: 0.000 min
Exp R.T. : 11.002 min
Response: 0
Conc: N.D.