

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061820\
 Data File : PS010851.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Jun 2020 01:31
 Operator : DD\AJ
 Sample : L3024-11
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 RO-7

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 19 02:51:52 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052920.M
 Quant Title : 8080.M
 QLast Update : Fri May 29 15:43:50 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.229	6.921	41136257	36700584	52.014	67.289 #
Target Compounds							
1) T	Dalapon	2.993	2.582	31213747	123.2E6	31.747	168.261 #
2) T	3,5-DICHL...	0.000	6.100	0	89960293	N.D.	109.755
5) T	DICAMBA	7.493	7.055	4795004	10341084	1.525	4.594 #
8) T	DICHLORPROP	0.000	7.740	0	13457660	N.D.	23.339
10) T	Pentachlo...	8.709	0.000	7299976	0	<MDL	N.D. #
12) T	2,4,5-T	9.561	0.000	8269903	0	2.033	N.D. #
14) T	DINOSEB	0.000	10.037	0	9513961	N.D.	4.422
15) T	Picloram	10.995	0.000	10875616	0	2.130	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061820\
Data File : PS010851.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Jun 2020 01:31
Operator : DD\AJ
Sample : L3024-11
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
RO-7

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 19 02:51:52 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052920.M
Quant Title : 8080.M
QLast Update : Fri May 29 15:43:50 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 2.993 min
Delta R.T.: -0.048 min
Response: 31213747
Conc: 31.75 ng/ml

Instrument :
ECD_S
ClientSampleId :
RO-7




#1 Dalapon

R.T.: 2.582 min
Delta R.T.: -0.041 min
Response: 123191676
Conc: 168.26 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 0.000 min
Exp R.T. : 6.443 min
Response: 0
Conc: N.D.



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 6.100 min
Delta R.T.: 0.030 min
Response: 89960293
Conc: 109.76 ng/ml



#4 2,4-DCAA


R.T.: 7.229 min
Delta R.T.: -0.002 min
Response: 41136257
Conc: 52.01 ng/ml

Instrument :
ECD_S
ClientSampleId :
RO-7




#4 2,4-DCAA

R.T.: 6.921 min
Delta R.T.: 0.000 min
Response: 36700584
Conc: 67.29 ng/ml




#5 DICAMBA

R.T.: 7.493 min
Delta R.T.: 0.070 min
Response: 4795004
Conc: 1.52 ng/ml



#5 DICAMBA


R.T.: 7.055 min
Delta R.T.: -0.036 min
Response: 10341084
Conc: 4.59 ng/ml



#8 DICHLORPROP


R.T.: 0.000 min
Exp R.T.: 8.116 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :
RO-7




#8 DICHLORPROP

R.T.: 7.740 min
Delta R.T.: 0.023 min
Response: 13457660
Conc: 23.34 ng/ml




#12 2,4,5-T

R.T.: 9.561 min
Delta R.T.: 0.069 min
Response: 8269903
Conc: 2.03 ng/ml



#12 2,4,5-T


R.T.: 0.000 min
Exp R.T.: 9.181 min
Response: 0
Conc: N.D.



#14 DINOSEB


R.T.: 0.000 min
Exp R.T.: 11.207 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :
RO-7




#14 DINOSEB

R.T.: 10.037 min
Delta R.T.: 0.015 min
Response: 9513961
Conc: 4.42 ng/ml



#15 Picloram

R.T.: 10.995 min
Delta R.T.: -0.029 min
Response: 10875616
Conc: 2.13 ng/ml



#15 Picloram

R.T.: 0.000 min
Exp R.T.: 11.002 min
Response: 0
Conc: N.D.