

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061821\
 Data File : PS015792.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Jun 2021 23:11
 Operator : DD\AJ
 Sample : M2672-13
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 19 02:11:44 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061521.M
 Quant Title : 8080.M
 QLast Update : Tue Jun 15 18:15:49 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml



System Monitoring Compounds							
4) S	2,4-DCAA	7.190	6.768	710.1E6	226.2E6	427.421	448.917
Target Compounds							
1) T	Dalapon	2.978f	2.483	29515643	50299991	14.624	77.446 #
2) T	3,5-DICHL...	6.442	5.983	-16220495	44723381	N.D.	59.302
3) T	4-Nitroph...	7.026	6.421	-23259417	11735107	N.D.	38.359
5) T	DICAMBA	7.362	6.890f	42152226	7349254	6.557	3.542 #
6) T	MCPD	7.549	0.000	19497702	0	3.590	N.D. #
7) T	MCPA	7.691	7.236	36085100	33450747	4.610	13.039 #
8) T	DICHLORPROP	8.092	7.570	35126911	26334761	22.185	50.864 #
9) T	2,4-D	8.317	7.827	461995	49004134	<MDL	79.002 #
10) T	Pentachlo...	8.632	8.332f	120.0E6	69610841	5.559	9.911 #
12) T	2,4,5-T	9.397f	9.021	60789878	19089162	6.886	6.768
13) T	2,4-DB	9.969	9.506	121.4E6	11132813	59.363	29.808 #
14) T	DINOSEB	11.120f	9.867	453.3E6	120.8E6	69.997	59.992
15) T	Picloram	10.947	10.834	292.1E6	74312945	21.364	17.670
16) T	DCPA	11.491f	10.834f	125.6E6	74312945	9.498	17.676 #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061821\
Data File : PS015792.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Jun 2021 23:11
Operator : DD\AJ
Sample : M2672-13
Misc :
ALS Vial : 26 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 19 02:11:44 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061521.M
Quant Title : 8080.M
QLast Update : Tue Jun 15 18:15:49 2021
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon

R.T.: 2.978 min
Delta R.T.: -0.042 min
Response: 29515643
Conc: 14.62 ng/ml




#1 Dalapon

R.T.: 2.483 min
Delta R.T.: -0.037 min
Response: 50299991
Conc: 77.45 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 6.442 min
Delta R.T.: 0.033 min
Response: -16220495
Conc: N.D.




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 5.983 min
Delta R.T.: 0.024 min
Response: 44723381
Conc: 59.30 ng/ml




#3 4-Nitrophenol

R.T.: 7.026 min
Delta R.T.: 0.007 min
Response: -23259417
Conc: N.D.




#3 4-Nitrophenol

R.T.: 6.421 min
Delta R.T.: 0.032 min
Response: 11735107
Conc: 38.36 ng/ml




#4 2,4-DCAA

R.T.: 7.190 min
Delta R.T.: -0.003 min
Response: 710136478
Conc: 427.42 ng/ml




#4 2,4-DCAA

R.T.: 6.768 min
Delta R.T.: -0.003 min
Response: 226187899
Conc: 448.92 ng/ml




#5 DICAMBA

R.T.: 7.362 min
Delta R.T.: -0.023 min
Response: 42152226
Conc: 6.56 ng/ml




#5 DICAMBA

R.T.: 6.890 min
Delta R.T.: -0.045 min
Response: 7349254
Conc: 3.54 ng/ml




#6 MCP

R.T.: 7.549 min
Delta R.T.: -0.012 min
Response: 19497702
Conc: 3.59 ug/ml




#6 MCP

R.T.: 0.000 min
Exp R.T.: 7.038 min
Response: 0
Conc: N.D.




#7 MCPA

R.T.: 7.691 min
Delta R.T.: -0.020 min
Response: 36085100
Conc: 4.61 ug/ml




#7 MCPA

R.T.: 7.236 min
Delta R.T.: -0.017 min
Response: 33450747
Conc: 13.04 ug/ml




#8 DICHLORPROP

R.T.: 8.092 min
Delta R.T.: 0.014 min
Response: 35126911
Conc: 22.19 ng/ml




#8 DICHLORPROP

R.T.: 7.570 min
Delta R.T.: 0.015 min
Response: 26334761
Conc: 50.86 ng/ml




#9 2,4-D

R.T.: 8.317 min
Delta R.T.: 0.011 min
Response: 461995
Conc: N.D.




#9 2,4-D

R.T.: 7.827 min
Delta R.T.: -0.018 min
Response: 49004134
Conc: 79.00 ng/ml




#10 Pentachlorophenol

R.T.: 8.632 min
Delta R.T.: 0.015 min
Response: 119975801
Conc: 5.56 ng/ml




#10 Pentachlorophenol

R.T.: 8.332 min
Delta R.T.: 0.057 min
Response: 69610841
Conc: 9.91 ng/ml




#12 2,4,5-T

R.T.: 9.397 min
Delta R.T.: -0.053 min
Response: 60789878
Conc: 6.89 ng/ml




#12 2,4,5-T

R.T.: 9.021 min
Delta R.T.: 0.008 min
Response: 19089162
Conc: 6.77 ng/ml




#13 2,4-DB

R.T.: 9.969 min
Delta R.T.: -0.031 min
Response: 121428870
Conc: 59.36 ng/ml




#13 2,4-DB

R.T.: 9.506 min
Delta R.T.: -0.013 min
Response: 11132813
Conc: 29.81 ng/ml




#14 DINOSEB

R.T.: 11.120 min
Delta R.T.: -0.043 min
Response: 453337620
Conc: 70.00 ng/ml




#14 DINOSEB

R.T.: 9.867 min
Delta R.T.: 0.020 min
Response: 120831096
Conc: 59.99 ng/ml




#15 Picloram

R.T.: 10.947 min
Delta R.T.: -0.031 min
Response: 292067384
Conc: 21.36 ng/ml




#15 Picloram

R.T.: 10.834 min
Delta R.T.: 0.010 min
Response: 74312945
Conc: 17.67 ng/ml



#16 DCPA

R.T.: 11.491 min
Delta R.T.: 0.060 min
Response: 125618163
Conc: 9.50 ng/ml



#16 DCPA

R.T.: 10.834 min
Delta R.T.: 0.055 min
Response: 74312945
Conc: 17.68 ng/ml