

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061821\
 Data File : PS015821.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Jun 2021 10:51
 Operator : DD\AJ
 Sample : M2637-09RE
 Misc :
 ALS Vial : 51 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 19 23:32:25 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061521.M
 Quant Title : 8080.M
 QLast Update : Tue Jun 15 18:15:49 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml



System Monitoring Compounds							
4) S	2,4-DCAA	7.188	6.766	974.1E6	273.7E6	586.276	543.140
Target Compounds							
1) T	Dalapon	2.976f	2.480f	21968803	29166185	10.884	44.907 #
2) T	3,5-DICHL...	6.368f	5.982	2256198	54679393	1.005	72.503 #
3) T	4-Nitroph...	7.027	0.000	2072742	0	2.280	N.D. #
5) T	DICAMBA	7.368	0.000	10295414	0	1.601	N.D. #
6) T	MCP	7.598	0.000	17330775	0	3.191	N.D. #
8) T	DICHLORPROP	0.000	7.569	0	17681333	N.D.	34.151
9) T	2,4-D	8.369f	0.000	33031946	0	18.848	N.D. #
10) T	Pentachlo...	8.676f	8.331f	12647102	8002238	<MDL	1.139 #
12) T	2,4,5-T	9.391f	9.068f	64553114	8903445	7.312	3.157 #
13) T	2,4-DB	10.038	0.000	18482765	0	9.036	N.D. #
14) T	DINOSEB	11.119f	9.878	198.1E6	252.3E6	30.594	125.285 #
15) T	Picloram	10.973	10.827	34994350	21910138	2.560	5.210 #
16) T	DCPA	0.000	10.827f	0	21910138	N.D.	5.211


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061821\
Data File : PS015821.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Jun 2021 10:51
Operator : DD\AJ
Sample : M2637-09RE
Misc :
ALS Vial : 51 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 19 23:32:25 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061521.M
Quant Title : 8080.M
QLast Update : Tue Jun 15 18:15:49 2021
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon

R.T.: 2.976 min
Delta R.T.: -0.044 min
Response: 21968803
Conc: 10.88 ng/ml




#1 Dalapon

R.T.: 2.480 min
Delta R.T.: -0.040 min
Response: 29166185
Conc: 44.91 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 6.368 min
Delta R.T.: -0.042 min
Response: 2256198
Conc: 1.01 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 5.982 min
Delta R.T.: 0.023 min
Response: 54679393
Conc: 72.50 ng/ml




#3 4-Nitrophenol

R.T.: 7.027 min
Delta R.T.: 0.008 min
Response: 2072742
Conc: 2.28 ng/ml




#3 4-Nitrophenol

R.T.: 0.000 min
Exp R.T. : 6.389 min
Response: 0
Conc: N.D.




#4 2,4-DCAA

R.T.: 7.188 min
Delta R.T.: -0.005 min
Response: 974066194
Conc: 586.28 ng/ml




#4 2,4-DCAA

R.T.: 6.766 min
Delta R.T.: -0.004 min
Response: 273662158
Conc: 543.14 ng/ml




#5 DICAMBA

R.T.: 7.368 min
Delta R.T.: -0.017 min
Response: 10295414
Conc: 1.60 ng/ml




#5 DICAMBA

R.T.: 0.000 min
Exp R.T. : 6.935 min
Response: 0
Conc: N.D.




#6 MCP

R.T.: 7.598 min
Delta R.T.: 0.037 min
Response: 17330775
Conc: 3.19 ug/ml




#6 MCP

R.T.: 0.000 min
Exp R.T. : 7.038 min
Response: 0
Conc: N.D.




#8 DICHLORPROP

R.T.: 0.000 min
Exp R.T. : 8.079 min
Response: 0
Conc: N.D.




#8 DICHLORPROP

R.T.: 7.569 min
Delta R.T.: 0.014 min
Response: 17681333
Conc: 34.15 ng/ml




#9 2,4-D

R.T.: 8.369 min
Delta R.T.: 0.063 min
Response: 33031946
Conc: 18.85 ng/ml




#9 2,4-D

R.T.: 0.000 min
Exp R.T. : 7.845 min
Response: 0
Conc: N.D.




#10 Pentachlorophenol

R.T.: 8.676 min
Delta R.T.: 0.058 min
Response: 12647102
Conc: N.D.




#10 Pentachlorophenol

R.T.: 8.331 min
Delta R.T.: 0.056 min
Response: 8002238
Conc: 1.14 ng/ml




#12 2,4,5-T

R.T.: 9.391 min
Delta R.T.: -0.059 min
Response: 64553114
Conc: 7.31 ng/ml




#12 2,4,5-T

R.T.: 9.068 min
Delta R.T.: 0.055 min
Response: 8903445
Conc: 3.16 ng/ml




#13 2,4-DB

R.T.: 10.038 min
Delta R.T.: 0.038 min
Response: 18482765
Conc: 9.04 ng/ml




#13 2,4-DB

R.T.: 0.000 min
Exp R.T. : 9.519 min
Response: 0
Conc: N.D.




#14 DINOSEB

R.T.: 11.119 min
Delta R.T.: -0.044 min
Response: 198138327
Conc: 30.59 ng/ml




#14 DINOSEB

R.T.: 9.878 min
Delta R.T.: 0.030 min
Response: 252341041
Conc: 125.29 ng/ml




#15 Picloram

R.T.: 10.973 min
Delta R.T.: -0.005 min
Response: 34994350
Conc: 2.56 ng/ml




#15 Picloram

R.T.: 10.827 min
Delta R.T.: 0.003 min
Response: 21910138
Conc: 5.21 ng/ml



#16 DCPA

R.T.: 0.000 min
Exp R.T. : 11.431 min
Response: 0
Conc: N.D.



#16 DCPA

R.T.: 10.827 min
Delta R.T.: 0.049 min
Response: 21910138
Conc: 5.21 ng/ml