

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061824\
 Data File : PS027010.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Jun 2024 09:19
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 18 11:56:03 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061724.M
 Quant Title : 8080.M
 QLast Update : Tue Jun 18 02:58:29 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.225	7.758	1836.8E6	1635.8E6	721.775	706.561
Target Compounds							
1) T	Dalapon	2.627	2.701	2310.5E6	2860.4E6	653.429	646.075
2) T	3,5-DICHL...	6.397	6.709	2727.0E6	2241.5E6	677.202	667.252
3) T	4-Nitroph...	7.024	7.285	1162.9E6	1085.0E6	657.986	653.025
5) T	DICAMBA	7.410	7.958	7353.1E6	6932.7E6	687.608	676.248
6) T	MCPD	7.589	8.059	520.1E6	495.4E6	71.894	67.465
7) T	MCPA	7.738	8.303	718.7E6	693.8E6	69.443	66.667
8) T	DICHLORPROP	8.109	8.671	1869.0E6	1764.5E6	679.199	667.186
9) T	2,4-D	8.338	9.001	2026.1E6	1954.4E6	681.773	679.820
10) T	Pentachlo...	8.632	9.531	27502.7E6	26600.0E6	714.037	704.280
11) T	2,4,5-TP ...	9.207	9.904	10455.7E6	10683.5E6	700.461	696.305
12) T	2,4,5-T	9.497	10.323	10590.2E6	10827.6E6	701.335	695.126
13) T	2,4-DB	10.067	10.888	1485.8E6	1366.1E6	696.083	687.147
14) T	DINOSEB	11.269	11.266	7403.6E6	7228.1E6	689.221	676.945
15) T	Picloram	11.081	12.361	14714.6E6	16302.6E6	700.029	694.928
16) T	DCPA	11.563	12.308	13549.3E6	12986.1E6	714.930	700.088

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061824\
Data File : PS027010.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Jun 2024 09:19
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 18 11:56:03 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061724.M
Quant Title : 8080.M
QLast Update : Tue Jun 18 02:58:29 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

