

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061918\
 Data File : PS000468.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Jun 2018 00:35
 Operator : UA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 20 01:16:38 2018
 Quant Method : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061718.M
 Quant Title : 8080.M
 QLast Update : Mon Jun 18 00:27:44 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.807	11.183	277.3E6	253.8E6	736.349	745.851
Target Compounds							
1) T	Dalapon	3.162	2.539	337.0E6	287.9E6	658.938	660.365
2) T	3,5-DICHL...	10.221	9.328	374.1E6	322.7E6	693.219	701.542
3) T	4-Nitroph...	11.405	10.330	159.0E6	133.8E6	676.577	682.938
5) T	DICAMBA	12.127	11.472	998.9E6	893.9E6	706.705	721.462
6) T	MCPD	12.465	11.698	78059640	69181359	68.932	72.105
7) T	MCPA	12.718	12.076	123.7E6	102.8E6	70.404	74.253
8) T	DICHLORPROP	13.319	12.595	262.0E6	227.8E6	696.425	707.497
9) T	2,4-D	13.677	13.067	339.1E6	307.7E6	708.891	724.018
10) T	Pentachlo...	14.075	13.631	3698.0E6	3435.2E6	704.387	726.165
11) T	2,4,5-TP ...	14.941	14.239	1536.6E6	1415.3E6	707.858	727.303
12) T	2,4,5-T	15.338	14.814	1519.9E6	1129.3E6	705.861	667.770
13) T	2,4-DB	16.092	15.458	203.3E6	180.6E6	690.467	729.365
14) T	DINOSEB	17.603	15.892	1022.0E6	909.9E6	686.342	697.507
15) T	Picloram	17.360	17.136	2054.2E6	1813.9E6	703.377	727.189
16) T	DCPA	17.942	17.078	1834.5E6	1697.6E6	715.898	736.184



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061918\
Data File : PS000468.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Jun 2018 00:35
Operator : UA
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 20 01:16:38 2018
Quant Method : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061718.M
Quant Title : 8080.M
QLast Update : Mon Jun 18 00:27:44 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 3.162 min
Delta R.T.: -0.020 min
Response: 337020737
Conc: 658.94 ng/ml

Instrument :
ECD_S
ClientSampleId :




#1 Dalapon

R.T.: 2.539 min
Delta R.T.: -0.021 min
Response: 287874849
Conc: 660.37 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 10.221 min
Delta R.T.: -0.022 min
Response: 374114092
Conc: 693.22 ng/ml



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 9.328 min
Delta R.T.: -0.020 min
Response: 322738335
Conc: 701.54 ng/ml



#3 4-Nitrophenol


R.T.: 11.405 min
Delta R.T.: -0.023 min
Response: 159043341
Conc: 676.58 ng/ml

Instrument :
ECD_S
ClientSampleId :




#3 4-Nitrophenol

R.T.: 10.330 min
Delta R.T.: -0.020 min
Response: 133829817
Conc: 682.94 ng/ml




#4 2,4-DCAA

R.T.: 11.807 min
Delta R.T.: -0.022 min
Response: 277349245
Conc: 736.35 ng/ml



#4 2,4-DCAA


R.T.: 11.183 min
Delta R.T.: -0.020 min
Response: 253819734
Conc: 745.85 ng/ml



#5 DICAMBA


R.T.: 12.127 min
Delta R.T.: -0.021 min
Response: 998893115
Conc: 706.71 ng/ml

Instrument :
ECD_S
ClientSampleId :




#5 DICAMBA

R.T.: 11.472 min
Delta R.T.: -0.020 min
Response: 893863009
Conc: 721.46 ng/ml




#6 MCP

R.T.: 12.465 min
Delta R.T.: -0.016 min
Response: 78059640
Conc: 68.93 ug/ml



#6 MCP


R.T.: 11.698 min
Delta R.T.: -0.013 min
Response: 69181359
Conc: 72.10 ug/ml



#7 MCPA


R.T.: 12.718 min
Delta R.T.: -0.015 min
Response: 123749200
Conc: 70.40 ug/ml

Instrument :
ECD_S
ClientSampleId :




#7 MCPA

R.T.: 12.076 min
Delta R.T.: -0.012 min
Response: 102836757
Conc: 74.25 ug/ml




#8 DICHLORPROP

R.T.: 13.319 min
Delta R.T.: -0.022 min
Response: 261966575
Conc: 696.43 ng/ml



#8 DICHLORPROP


R.T.: 12.595 min
Delta R.T.: -0.020 min
Response: 227793921
Conc: 707.50 ng/ml



#9 2,4-D


R.T.: 13.677 min
Delta R.T.: -0.022 min
Response: 339065613
Conc: 708.89 ng/ml

Instrument :
ECD_S
ClientSampleId :




#9 2,4-D

R.T.: 13.067 min
Delta R.T.: -0.020 min
Response: 307707395
Conc: 724.02 ng/ml




#10 Pentachlorophenol

R.T.: 14.075 min
Delta R.T.: -0.021 min
Response: 3698007508
Conc: 704.39 ng/ml



#10 Pentachlorophenol


R.T.: 13.631 min
Delta R.T.: -0.020 min
Response: 3435245728
Conc: 726.16 ng/ml



#11 2,4,5-TP (SILVEX)


R.T.: 14.941 min
Delta R.T.: -0.021 min
Response: 1536646366
Conc: 707.86 ng/ml

Instrument :
ECD_S
ClientSampleId :




#11 2,4,5-TP (SILVEX)

R.T.: 14.239 min
Delta R.T.: -0.020 min
Response: 1415342284
Conc: 727.30 ng/ml




#12 2,4,5-T

R.T.: 15.338 min
Delta R.T.: -0.022 min
Response: 1519914925
Conc: 705.86 ng/ml



#12 2,4,5-T


R.T.: 14.814 min
Delta R.T.: -0.020 min
Response: 1129287291
Conc: 667.77 ng/ml



#13 2,4-DB


R.T.: 16.092 min
Delta R.T.: -0.022 min
Response: 203286124
Conc: 690.47 ng/ml

Instrument :
ECD_S
ClientSampleId :




#13 2,4-DB

R.T.: 15.458 min
Delta R.T.: -0.020 min
Response: 180552359
Conc: 729.36 ng/ml




#14 DINOSEB

R.T.: 17.603 min
Delta R.T.: -0.022 min
Response: 1022048032
Conc: 686.34 ng/ml



#14 DINOSEB


R.T.: 15.892 min
Delta R.T.: -0.020 min
Response: 909946626
Conc: 697.51 ng/ml



#15 Picloram


R.T.: 17.360 min
Delta R.T.: -0.022 min
Response: 2054197389
Conc: 703.38 ng/ml

Instrument :
ECD_S
ClientSampleId :




#15 Picloram

R.T.: 17.136 min
Delta R.T.: -0.020 min
Response: 1813867962
Conc: 727.19 ng/ml



#16 DCPA

R.T.: 17.942 min
Delta R.T.: -0.022 min
Response: 1834468164
Conc: 715.90 ng/ml



#16 DCPA

R.T.: 17.078 min
Delta R.T.: -0.020 min
Response: 1697612875
Conc: 736.18 ng/ml