

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\
 Method File : PS061919.M
 Title : 8080.M
 Last Update : Thu Jun 20 11:04:04 2019
 Response Via : Initial Calibration

Calibration Files

750 =PS005119.D 1000 =PS005120.D 1500 =PS005121.D
 500 =PS005118.D 200 =PS005117.D

	Compound	750	1000	1500	500	200	Avg		%RSD
1) T	Dalapon	1.824	1.891	1.865	1.890	1.747	1.843	E6	3.28
2) T	3,5-DICHLOROBENZOI...	1.970	1.996	1.936	2.089	2.058	2.010	E6	3.11
3) T	4-Nitrophenol	7.123	7.393	7.519	7.358	6.822	7.243	E5	3.80
4) S	2,4-DCAA	1.441	1.475	1.467	1.508	1.449	1.468	E6	1.77
5) T	DICAMBA	5.627	5.669	5.448	5.979	5.801	5.705	E6	3.48
6) T	MCPP	4.378	4.635	4.780	4.434	3.660	4.377	E6	9.87
7) T	MCPA	6.191	6.496	6.769	6.321	5.622	6.280	E6	6.79
8) T	DICHLORPROP	1.407	1.440	1.367	1.475	1.461	1.430	E6	3.03
9) T	2,4-D	1.525	1.563	1.549	1.600	1.536	1.555	E6	1.89
10) T	Pentachlorophenol	1.929	1.864	1.692	2.133	2.220	1.968	E7	10.77
11) T	2,4,5-TP (SILVEX)	7.997	7.990	7.601	8.604	8.599	8.158	E6	5.33
12) T	2,4,5-T	7.904	7.907	7.596	8.471	8.361	8.048	E6	4.49
13) T	2,4-DB	1.142	1.180	1.196	1.184	1.105	1.162	E6	3.22
14) T	DINOSEB	6.327	6.348	6.129	6.629	6.525	6.392	E6	3.02
15) T	Picloram	1.143	1.170	1.133	1.193	1.106	1.149	E7	2.91
16) T	DCPA	0.983	0.972	0.916	1.064	1.053	0.998	E7	6.13

Signal #2 Calibration Files

750 =PS005119.D 1000 =PS005120.D 1500 =PS005121.D
 500 =PS005118.D 200 =PS005117.D

	Compound	750	1000	1500	500	200	Avg		%RSD
1) T	Dalapon	1.316	1.370	1.396	1.283	1.230	1.319	E6	5.05
2) T	3,5-DICHLOROBENZOI...	1.296	1.312	1.273	1.374	1.338	1.319	E6	2.96
3) T	4-Nitrophenol	4.906	5.026	4.986	5.149	5.005	5.014	E5	1.75
4) S	2,4-DCAA	0.978	1.006	0.994	1.026	1.009	1.002	E6	1.78
5) T	DICAMBA	3.669	3.700	3.546	3.918	3.884	3.743	E6	4.16
6) T	MCPP	2.772	2.911	2.952	2.826	2.468	2.785	E6	6.87
7) T	MCPA	3.970	4.129	4.127	4.114	3.915	4.051	E6	2.50
8) T	DICHLORPROP	9.420	9.684	9.632	9.881	9.660	9.655	E5	1.70
9) T	2,4-D	1.088	1.115	1.105	1.146	1.123	1.115	E6	1.91
10) T	Pentachlorophenol	1.254	1.207	1.085	1.401	1.481	1.286	E7	12.25
11) T	2,4,5-TP (SILVEX)	5.197	5.225	4.972	5.593	5.639	5.325	E6	5.32
12) T	2,4,5-T	5.029	5.033	4.774	5.377	5.292	5.101	E6	4.69
13) T	2,4-DB	6.459	6.786	6.895	6.593	5.910	6.529	E5	5.90
14) T	DINOSEB	3.960	4.015	3.901	4.199	4.142	4.043	E6	3.07
15) T	Picloram	7.483	7.631	7.332	7.871	7.415	7.546	E6	2.81
16) T	DCPA	6.684	6.589	6.162	7.271	7.331	6.807	E6	7.22

(#) = Out of Range