

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061919\
 Data File : PS005117.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Jun 2019 10:33
 Operator : SM\AJ
 Sample : HSTDICC200
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC200

Manual Integrations
 APPROVED

Ankita
 6/20/2019 2:02:29 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 19 12:46:04 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061919.M
 Quant Title : 8080.M
 QLast Update : Wed Jun 19 12:45:35 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.611	5.303	289.8E6	201.8E6	201.060	206.434
Target Compounds							
1) T	Dalapon	2.138	1.811	317.9E6	223.9E6	174.302m	170.181
2) T	3,5-DICHL...	4.961	4.596	382.7E6	248.8E6	194.237	192.000
3) T	4-Nitroph...	5.456	4.978	124.2E6	91098826	174.313	185.685
5) T	DICAMBA	5.771	5.436	1090.6E6	730.3E6	193.838	199.065
6) T	MCPD	5.926	5.521	68806481	46390824	15.715	16.738
7) T	MCPA	6.062	5.704	104.6E6	72813459	16.891	18.342
8) T	DICHLORPROP	6.409	5.992	274.6E6	181.6E6	195.136	192.807
9) T	2,4-D	6.624	6.264	288.8E6	211.1E6	189.407	193.938
10) T	Pentachlo...	6.897	6.641	4218.6E6	2814.8E6	218.656	224.393
11) T	2,4,5-TP ...	7.433	7.004	1633.9E6	1071.4E6	204.300	206.148
12) T	2,4,5-T	7.706	7.365	1588.6E6	1005.4E6	200.971	199.933
13) T	2,4-DB	8.238	7.854	210.0E6	112.3E6	183.848	173.843
14) T	DINOSEB	9.340	8.169	1226.7E6	778.6E6	193.892	196.613
15) T	Picloram	9.167	9.110	2102.3E6	1408.8E6	183.858	188.277
16) T	DCPA	9.611	9.070	2020.8E6	1407.5E6	205.499	210.567

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061919\
Data File : PS005117.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Jun 2019 10:33
Operator : SM\AJ
Sample : HSTDICC200
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC200

Manual Integrations
APPROVED

Ankita
6/20/2019 2:02:29 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 19 12:46:04 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061919.M
Quant Title : 8080.M
QLast Update : Wed Jun 19 12:45:35 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

