

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061919\
 Data File : PS005122.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Jun 2019 12:23
 Operator : SM\AJ
 Sample : HSTDICV750
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 ICVPS061919

Manual Integrations
 APPROVED

Ankita
 6/20/2019 2:02:33 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 20 11:20:46 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061919.M
 Quant Title : 8080.M
 QLast Update : Thu Jun 20 11:04:04 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.571	5.286	1094.8E6	763.0E6	745.860	761.244
Target Compounds							
1) T	Dalapon	2.098	1.795	1318.2E6	963.8E6	715.121m	730.664m
2) T	3,5-DICHL...	4.921	4.579	1396.3E6	928.2E6	694.769	703.836
3) T	4-Nitroph...	5.414	4.960	486.2E6	347.9E6	671.251m	693.695
5) T	DICAMBA	5.732	5.420	4115.5E6	2697.7E6	721.426	720.675
6) T	MCPD	5.892	5.510	324.2E6	205.6E6	74.050	73.809
7) T	MCPA	6.030	5.695	413.3E6	289.6E6	65.821	71.496
8) T	DICHLORPROP	6.370	5.975	1000.5E6	689.4E6	699.654	714.019
9) T	2,4-D	6.583	6.247	1116.5E6	798.0E6	718.202	715.513
10) T	Pentachlo...	6.857	6.625	14131.8E6	9182.4E6	718.222	714.231
11) T	2,4,5-TP ...	7.393	6.987	5900.4E6	3830.3E6	723.247	719.271
12) T	2,4,5-T	7.666	7.348	5853.7E6	3698.9E6	727.368	725.145
13) T	2,4-DB	8.197	7.836	837.1E6	485.8E6	720.707	744.129
14) T	DINOSEB	9.301	8.152	4637.0E6	2916.0E6	725.465	721.203
15) T	Picloram	9.124	9.092	8553.3E6	5607.5E6	744.339	743.062
16) T	DCPA	9.572	9.053	7312.2E6	4981.0E6	733.032	731.706

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061919\
Data File : PS005122.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Jun 2019 12:23
Operator : SM\AJ
Sample : HSTDICV750
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
ICVPS061919

Manual Integrations
APPROVED

Ankita
6/20/2019 2:02:33 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 20 11:20:46 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061919.M
Quant Title : 8080.M
QLast Update : Thu Jun 20 11:04:04 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

