

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061919\
 Data File : PS005124.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Jun 2019 13:21
 Operator : SM\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Ankita
 6/20/2019 2:02:35 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 20 11:16:33 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061919.M
 Quant Title : 8080.M
 QLast Update : Thu Jun 20 11:04:04 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.585	5.283	1068.4E6	725.8E6	727.856	724.130
Target Compounds							
1) T	Dalapon	2.114	1.794	1246.0E6	933.7E6	675.941	707.859m
2) T	3,5-DICHL...	4.935	4.576	1343.1E6	884.3E6	668.283	670.570
3) T	4-Nitroph...	5.429	4.957	481.6E6	334.6E6	664.861	667.216
5) T	DICAMBA	5.746	5.417	3967.0E6	2594.9E6	695.381	693.216
6) T	MCPD	5.906	5.506	311.7E6	196.5E6	71.209	70.543
7) T	MCPA	6.044	5.691	436.5E6	279.4E6	69.510	68.963
8) T	DICHLORPROP	6.385	5.972	985.6E6	661.7E6	689.210	685.271
9) T	2,4-D	6.599	6.243	1069.1E6	765.3E6	687.709	686.175
10) T	Pentachlo...	6.872	6.622	13610.6E6	8848.9E6	691.733	688.297
11) T	2,4,5-TP ...	7.407	6.984	5668.3E6	3677.5E6	694.792	690.573
12) T	2,4,5-T	7.680	7.345	5622.8E6	3546.5E6	698.676	695.287
13) T	2,4-DB	8.212	7.833	814.0E6	470.0E6	700.750	719.960
14) T	DINOSEB	9.315	8.149	4461.8E6	2801.6E6	698.057	692.909
15) T	Picloram	9.139	9.089	8406.2E6	5514.9E6	731.537	730.799
16) T	DCPA	9.586	9.050	7024.4E6	4775.4E6	704.182	701.495

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061919\
Data File : PS005124.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Jun 2019 13:21
Operator : SM\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
HSTDCCC750

Manual Integrations
APPROVED

Ankita
6/20/2019 2:02:35 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 20 11:16:33 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061919.M
Quant Title : 8080.M
QLast Update : Thu Jun 20 11:04:04 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

