

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061919\
 Data File : PS005131.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Jun 2019 16:39
 Operator : SM\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Ankita
 6/20/2019 2:02:46 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 20 11:17:50 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061919.M
 Quant Title : 8080.M
 QLast Update : Thu Jun 20 11:04:04 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.587	5.302	1055.4E6	729.1E6	718.956	727.350
Target Compounds							
1) T	Dalapon	2.113	1.812	1230.4E6	936.7E6	667.511	710.098m
2) T	3,5-DICHL...	4.937	4.595	1325.1E6	873.3E6	659.336	662.229
3) T	4-Nitroph...	5.431	4.976	471.1E6	332.6E6	650.479	663.341
5) T	DICAMBA	5.748	5.436	3971.0E6	2607.3E6	696.084	696.531
6) T	MCPD	5.908	5.525	314.2E6	195.9E6	71.787	70.338
7) T	MCPA	6.046	5.710	439.6E6	281.7E6	70.010	69.533
8) T	DICHLORPROP	6.386	5.991	973.8E6	661.6E6	680.962m	685.192
9) T	2,4-D	6.601	6.263	1057.4E6	760.3E6	680.198	681.719
10) T	Pentachlo...	6.874	6.641	13504.4E6	8812.4E6	686.332	685.454
11) T	2,4,5-TP ...	7.410	7.002	5612.3E6	3679.2E6	687.935	690.897
12) T	2,4,5-T	7.683	7.364	5586.6E6	3494.5E6	694.169	685.085
13) T	2,4-DB	8.215	7.852	802.7E6	468.3E6	691.073	717.361
14) T	DINOSEB	9.317	8.168	4437.4E6	2746.0E6	694.251	679.155
15) T	Picloram	9.142	9.069	7835.3E6	4670.8E6	681.852	618.941
16) T	DCPA	9.588	9.069	6894.7E6	4670.8E6	691.182	686.131

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061919\
Data File : PS005131.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Jun 2019 16:39
Operator : SM\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
HSTDCCC750

Manual Integrations
APPROVED

Ankita
6/20/2019 2:02:46 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 20 11:17:50 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061919.M
Quant Title : 8080.M
QLast Update : Thu Jun 20 11:04:04 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

