

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061920\
 Data File : PS010873.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Jun 2020 19:34
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 20 03:48:15 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052920.M
 Quant Title : 8080.M
 QLast Update : Fri May 29 15:43:50 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.228	6.920	699.3E6	447.9E6	884.286	821.134
Target Compounds							
1) T	Dalapon	3.041	2.625	733.0E6	486.2E6	745.481	664.106
2) T	3,5-DICHL...	6.441	6.070	905.2E6	618.6E6	828.030	754.664
3) T	4-Nitroph...	7.058	6.525	289.0E6	236.8E6	779.607	714.748
5) T	DICAMBA	7.420	7.089	2515.4E6	1689.6E6	799.941	750.660
6) T	MCPD	7.590	7.187	202.9E6	141.4E6	82.121	75.899
7) T	MCPA	7.739	7.403	251.8E6	199.3E6	72.015	73.320
8) T	DICHLORPROP	8.112	7.714	613.2E6	439.8E6	809.073	762.759
9) T	2,4-D	8.341	8.008	633.8E6	546.6E6	804.034	720.851
10) T	Pentachlo...	8.653	8.439	8957.6E6	5334.6E6	839.821	750.878
11) T	2,4,5-TP ...	9.202	8.799	3541.5E6	2319.0E6	828.075	772.144
12) T	2,4,5-T	9.487	9.178	3323.6E6	2287.1E6	817.200	762.139
13) T	2,4-DB	10.040	9.685	402.2E6	308.8E6	727.343	760.140
14) T	DINOSEB	11.202	10.019	2646.5E6	1693.4E6	819.060	787.109
15) T	Picloram	11.021	11.001	2976.8E6	2338.2E6	583.098	566.699
16) T	DCPA	11.467	10.948	4412.5E6	2928.8E6	823.532	789.807

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061920\
Data File : PS010873.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Jun 2020 19:34
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 20 03:48:15 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052920.M
Quant Title : 8080.M
QLast Update : Fri May 29 15:43:50 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

