

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061921\
 Data File : PS015864.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Jun 2021 05:00
 Operator : DD\AJ
 Sample : M2639-01
 Misc :
 ALS Vial : 65 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 21 01:51:33 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061521.M
 Quant Title : 8080.M
 QLast Update : Tue Jun 15 18:15:49 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.190	6.767	113.6E6	26045551	68.352	51.693
Target Compounds							
1) T	Dalapon	2.974f	2.481	20425637	22789470	10.120	35.089 #
2) T	3,5-DICHL...	0.000	5.983	0	67696944	N.D.	89.764
5) T	DICAMBA	7.367	0.000	16022885	0	2.492	N.D. #
6) T	MCP	7.547	0.000	17549666	0	3.231	N.D. #
7) T	MCPA	7.675	7.226	15450494	19390035	1.974	7.558 #
8) T	DICHLORPROP	8.025f	7.569	13225945	13030613	8.353	25.168 #
9) T	2,4-D	8.306	7.873	7690407	29688409	4.388	47.862 #
10) T	Pentachlo...	8.610	8.269	5878282	8259889	<MDL	1.176 #
12) T	2,4,5-T	9.512f	0.000	2875394	0	<MDL	N.D. #
13) T	2,4-DB	9.945f	0.000	41947729	0	20.507	N.D. #
14) T	DINOSEB	11.121f	9.860	111.7E6	20429059	17.248	10.143 #
15) T	Picloram	10.972	10.819	107.0E6	24311560	7.829	5.781 #
16) T	DCPA	0.000	10.819f	0	24311560	N.D.	5.783



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061921\
Data File : PS015864.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Jun 2021 05:00
Operator : DD\AJ
Sample : M2639-01
Misc :
ALS Vial : 65 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 21 01:51:33 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061521.M
Quant Title : 8080.M
QLast Update : Tue Jun 15 18:15:49 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 2.974 min
Delta R.T.: -0.046 min
Response: 20425637
Conc: 10.12 ng/ml

Instrument :
ECD_S
ClientSampleId :




#1 Dalapon

R.T.: 2.481 min
Delta R.T.: -0.040 min
Response: 22789470
Conc: 35.09 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 0.000 min
Exp R.T. : 6.410 min
Response: 0
Conc: N.D.



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 5.983 min
Delta R.T.: 0.023 min
Response: 67696944
Conc: 89.76 ng/ml



#4 2,4-DCAA


R.T.: 7.190 min
Delta R.T.: -0.003 min
Response: 113562810
Conc: 68.35 ng/ml

Instrument :
ECD_S
ClientSampleId :




#4 2,4-DCAA

R.T.: 6.767 min
Delta R.T.: -0.004 min
Response: 26045551
Conc: 51.69 ng/ml




#5 DICAMBA

R.T.: 7.367 min
Delta R.T.: -0.019 min
Response: 16022885
Conc: 2.49 ng/ml



#5 DICAMBA


R.T.: 0.000 min
Exp R.T.: 6.935 min
Response: 0
Conc: N.D.



#6 MCPP


R.T.: 7.547 min
Delta R.T.: -0.014 min
Response: 17549666
Conc: 3.23 ug/ml

Instrument :
ECD_S
ClientSampleId :




#6 MCPP

R.T.: 0.000 min
Exp R.T. : 7.038 min
Response: 0
Conc: N.D.




#7 MCPA

R.T.: 7.675 min
Delta R.T.: -0.036 min
Response: 15450494
Conc: 1.97 ug/ml



#7 MCPA


R.T.: 7.226 min
Delta R.T.: -0.026 min
Response: 19390035
Conc: 7.56 ug/ml



#8 DICHLORPROP


R.T.: 8.025 min
Delta R.T.: -0.054 min
Response: 13225945
Conc: 8.35 ng/ml

Instrument :
ECD_S
ClientSampleId :




#8 DICHLORPROP

R.T.: 7.569 min
Delta R.T.: 0.014 min
Response: 13030613
Conc: 25.17 ng/ml




#9 2,4-D

R.T.: 8.306 min
Delta R.T.: 0.000 min
Response: 7690407
Conc: 4.39 ng/ml



#9 2,4-D


R.T.: 7.873 min
Delta R.T.: 0.027 min
Response: 29688409
Conc: 47.86 ng/ml



#10 Pentachlorophenol


R.T.: 8.610 min
Delta R.T.: -0.008 min
Response: 5878282
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#10 Pentachlorophenol

R.T.: 8.269 min
Delta R.T.: -0.007 min
Response: 8259889
Conc: 1.18 ng/ml




#11 2,4,5-TP (SILVEX)

R.T.: 9.195 min
Delta R.T.: 0.027 min
Response: 32423444
Conc: 3.71 ng/ml



#11 2,4,5-TP (SILVEX)


R.T.: 8.659 min
Delta R.T.: 0.023 min
Response: -75820287
Conc: N.D.



#13 2,4-DB


R.T.: 9.945 min
Delta R.T.: -0.055 min
Response: 41947729
Conc: 20.51 ng/ml

Instrument :
ECD_S
ClientSampleId :




#13 2,4-DB

R.T.: 0.000 min
Exp R.T. : 9.519 min
Response: 0
Conc: N.D.




#14 DINOSEB

R.T.: 11.121 min
Delta R.T.: -0.043 min
Response: 111708951
Conc: 17.25 ng/ml



#14 DINOSEB


R.T.: 9.860 min
Delta R.T.: 0.013 min
Response: 20429059
Conc: 10.14 ng/ml



#15 Picloram


R.T.: 10.972 min
Delta R.T.: -0.006 min
Response: 107026551
Conc: 7.83 ng/ml

Instrument :
ECD_S
ClientSampleId :




#15 Picloram

R.T.: 10.819 min
Delta R.T.: -0.005 min
Response: 24311560
Conc: 5.78 ng/ml



#16 DCPA

R.T.: 0.000 min
Exp R.T. : 11.431 min
Response: 0
Conc: N.D.



#16 DCPA

R.T.: 10.819 min
Delta R.T.: 0.041 min
Response: 24311560
Conc: 5.78 ng/ml