

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061921\
 Data File : PS015851.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Jun 2021 22:57
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 115 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 20 00:56:03 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061521.M
 Quant Title : 8080.M
 QLast Update : Tue Jun 15 18:15:49 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.189	6.767	1312.5E6	416.2E6	789.976	825.955
Target Compounds							
1) T	Dalapon	3.017	2.517	1568.5E6	464.9E6	777.096	715.852
2) T	3,5-DICHL...	6.407	5.958	1727.9E6	585.9E6	769.853	776.940
3) T	4-Nitroph...	7.016	6.387	673.2E6	228.6E6	740.421	747.199
5) T	DICAMBA	7.381	6.931	4673.5E6	1640.4E6	726.951	790.553
6) T	MCPD	7.552	7.028	365.7E6	132.1E6	67.327	79.256
7) T	MCPA	7.701	7.241	514.7E6	193.6E6	65.743	75.482
8) T	DICHLORPROP	8.074	7.551	1117.0E6	402.4E6	705.444	777.276
9) T	2,4-D	8.301	7.841	1223.5E6	479.9E6	698.100	773.652
10) T	Pentachlo...	8.612	8.269	15465.7E6	5465.2E6	716.557	778.114
11) T	2,4,5-TP ...	9.162	8.630	5901.3E6	2190.2E6	675.108	793.333
12) T	2,4,5-T	9.445	9.007	5981.0E6	2227.4E6	677.479	789.740
13) T	2,4-DB	9.995	9.513	1293.1E6	294.6E6	632.142	788.846
14) T	DINOSEB	11.156	9.840	4190.4E6	1574.9E6	647.012	781.913
15) T	Picloram	10.973	10.817	7647.9E6	2985.8E6	559.433	709.957 #
16) T	DCPA	11.424	10.771	8364.8E6	3348.7E6	632.428	796.520 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061921\
Data File : PS015851.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Jun 2021 22:57
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 115 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 20 00:56:03 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061521.M
Quant Title : 8080.M
QLast Update : Tue Jun 15 18:15:49 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

