

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061923\
 Data File : PS023326.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Jun 2023 16:07
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 20 01:34:21 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS060623.M
 Quant Title : 8080.M
 QLast Update : Tue Jun 06 13:00:04 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.480	6.857	1310.1E6	423.9E6	740.480	746.672
Target Compounds							
1) T	Dalapon	2.550	2.576	1252.2E6	623.6E6	684.806	680.356
2) T	3,5-DICHL...	5.833	6.023	1862.5E6	620.8E6	693.257	701.290
3) T	4-Nitroph...	6.312	6.461	884.1E6	226.9E6	685.979	665.211
5) T	DICAMBA	6.628	7.023	5543.1E6	1841.8E6	711.225	718.695
6) T	MCPD	6.787	7.118	402.9E6	128.8E6	73.098	72.047
7) T	MCPA	6.915	7.329	553.4E6	192.3E6	70.674	71.560
8) T	DICHLORPROP	7.237	7.645	1426.0E6	435.7E6	691.058	704.349
9) T	2,4-D	7.438	7.932	1692.5E6	489.7E6	692.692	706.170
10) T	Pentachlo...	7.673	8.363	20126.6E6	6571.0E6	714.231	726.327
11) T	2,4,5-TP ...	8.207	8.723	8197.3E6	2421.0E6	708.917	725.437
12) T	2,4,5-T	8.468	9.096	8006.2E6	2196.3E6	704.757	716.431
13) T	2,4-DB	8.982	9.602	1579.1E6	270.0E6	670.776	728.199
14) T	DINOSEB	10.055	9.932	5807.7E6	1705.2E6	701.224	718.579
15) T	Picloram	9.876	10.896	11168.1E6	2874.8E6	643.525	629.915
16) T	DCPA	10.324	10.859	9983.2E6	2939.1E6	716.131	732.592

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061923\
Data File : PS023326.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Jun 2023 16:07
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 20 01:34:21 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS060623.M
Quant Title : 8080.M
QLast Update : Tue Jun 06 13:00:04 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

