

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS062019\
 Data File : PS005141.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Jun 2019 08:25
 Operator : SM\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Ankita
 6/21/2019 2:31:46 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 20 12:00:47 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061919.M
 Quant Title : 8080.M
 QLast Update : Thu Jun 20 11:04:04 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.569	5.283	1048.5E6	756.7E6	714.297	754.876
Target Compounds							
1) T	Dalapon	2.096	1.793	1279.8E6	962.8E6	694.291m	729.907m
2) T	3,5-DICHL...	4.919	4.576	1359.8E6	900.7E6	676.609	682.979
3) T	4-Nitroph...	5.413	4.958	459.0E6	346.5E6	633.731m	690.971
5) T	DICAMBA	5.729	5.417	4109.3E6	2713.0E6	720.325	724.765
6) T	MCPD	5.890	5.508	328.0E6	208.5E6	74.930	74.842
7) T	MCPA	6.027	5.693	391.1E6	293.3E6	62.280	72.403
8) T	DICHLORPROP	6.368	5.972	985.2E6	692.1E6	688.948	716.785
9) T	2,4-D	6.582	6.244	1096.1E6	794.2E6	705.082	712.141
10) T	Pentachlo...	6.856	6.622	13828.4E6	9047.5E6	702.799	703.741
11) T	2,4,5-TP ...	7.391	6.984	5801.3E6	3810.3E6	711.094	715.502
12) T	2,4,5-T	7.664	7.345	5773.8E6	3630.3E6	717.440	711.705
13) T	2,4-DB	8.196	7.833	820.4E6	492.7E6	706.261	754.762
14) T	DINOSEB	9.298	8.149	4514.1E6	2833.0E6	706.241	700.685
15) T	Picloram	9.123	9.089	8359.9E6	5462.0E6	727.509	723.784
16) T	DCPA	9.570	9.049	7081.1E6	4841.9E6	709.869	711.273

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS062019\
Data File : PS005141.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Jun 2019 08:25
Operator : SM\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Ankita
6/21/2019 2:31:46 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 20 12:00:47 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061919.M
Quant Title : 8080.M
QLast Update : Thu Jun 20 11:04:04 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

