

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS062024\
 Data File : PS027063.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Jun 2024 11:29
 Operator : AR\AJ
 Sample : PB161593BSD
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PB161593BSD

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 06/21/2024
 Supervised By : Ankita Jodhani 06/25/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 21 01:46:09 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061724.M
 Quant Title : 8080.M
 QLast Update : Tue Jun 18 02:58:29 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.217	7.755	1346.1E6	1162.2E6	528.936	501.996
Target Compounds							
1) T	Dalapon	2.626	2.702	1609.4E6	2000.4E6	455.147	451.819
2) T	3,5-DICHL...	6.391	6.706	1975.8E6	1583.2E6	490.650	471.273
3) T	4-Nitroph...	7.018	7.282	826.5E6	745.1E6	467.655	448.453
5) T	DICAMBA	7.402	7.954	5289.7E6	4839.7E6	494.657	472.083
6) T	MCPD	7.580	8.054	356.3E6	339.2E6	49.256m	46.194
7) T	MCPA	7.729	8.298	488.7E6	475.4E6	47.220	45.681
8) T	DICHLORPROP	8.101	8.668	1362.3E6	1246.7E6	495.086	471.385
9) T	2,4-D	8.330	8.998	1427.1E6	1370.5E6	480.209	476.720
10) T	Pentachlo...	8.623	9.526	19897.9E6	18856.5E6	516.597	499.258
11) T	2,4,5-TP ...	9.198	9.900	7559.6E6	7501.2E6	506.442	488.895
12) T	2,4,5-T	9.489	10.319	7616.3E6	7598.0E6	504.393	487.784
13) T	2,4-DB	10.059	10.884	1057.1E6	955.6E6	495.213	480.642
14) T	DINOSEB	11.258	11.263	5451.6E6	5148.2E6	507.503	482.152
15) T	Picloram	11.070	12.358	10201.2E6	11150.7E6	485.312	475.321
16) T	DCPA	11.553	12.305	9925.1E6	9212.4E6	523.702	496.644

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS062024\
Data File : PS027063.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Jun 2024 11:29
Operator : AR\AJ
Sample : PB161593BSD
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
PB161593BSD

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 06/21/2024
Supervised By : Ankita Jodhani 06/25/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 21 01:46:09 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061724.M
Quant Title : 8080.M
QLast Update : Tue Jun 18 02:58:29 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

