

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS062121\
 Data File : PS015900.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Jun 2021 15:35
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 22 03:56:11 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061521.M
 Quant Title : 8080.M
 QLast Update : Tue Jun 15 18:15:49 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Ankita
 6/22/2021 11:37:46 AM

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.188	6.766	1303.8E6	424.9E6	784.730	843.324
Target Compounds							
1) T	Dalapon	3.015	2.517	1538.9E6	457.0E6	762.469	703.658
2) T	3,5-DICHL...	6.406	5.957	1711.9E6	596.8E6	762.725	791.357
3) T	4-Nitroph...	7.016	6.386	678.0E6	233.2E6	745.797	762.371
5) T	DICAMBA	7.380	6.929	4662.3E6	1670.9E6	725.196m	805.257
6) T	MCP	7.551	7.027	365.1E6	133.1E6	67.218	79.828
7) T	MCPA	7.700	7.239	512.0E6	197.4E6	65.407	76.965
8) T	DICHLORPROP	8.073	7.550	1112.2E6	408.8E6	702.460	789.538
9) T	2,4-D	8.301	7.840	1209.1E6	488.4E6	689.925	787.388
10) T	Pentachlo...	8.611	8.268	15378.5E6	5541.0E6	712.514	788.898
11) T	2,4,5-TP ...	9.162	8.629	5883.7E6	2229.5E6	673.098	807.564
12) T	2,4,5-T	9.445	9.007	5963.1E6	2257.9E6	675.455	800.542
13) T	2,4-DB	9.994	9.512	1260.7E6	298.4E6	616.339	798.847 #
14) T	DINOSEB	11.156	9.840	4185.9E6	1590.2E6	646.319	789.528
15) T	Picloram	10.972	10.816	8121.9E6	3211.4E6	594.101	763.599 #
16) T	DCPA	11.423	10.770	8405.1E6	3404.6E6	635.476	809.797 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061212\
Data File : PS015900.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Jun 2021 15:35
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 22 03:56:11 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061521.M
Quant Title : 8080.M
QLast Update : Tue Jun 15 18:15:49 2021
Response via : Initial Calibration
Integrator: ChemStation

Instrument :
ECD_S
ClientSampled :
HSTDCCC750

Manual Integrations
APPROVED

Ankita
6/22/2021 11:37:46 AM

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

