

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS062218\
 Data File : PS000573.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Jun 2018 16:09
 Operator : UA
 Sample : J3603-18
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 22 17:47:13 2018
 Quant Method : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061718.M
 Quant Title : 8080.M
 QLast Update : Mon Jun 18 00:27:44 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.802	11.180	128.2E6	105.2E6	340.431	309.273
Target Compounds							
2) T	3,5-DICHL...	0.000	9.328	0	763677	N.D.	1.660 #
3) T	4-Nitroph...	11.460	10.338	4000969	6519143	17.020	33.267 #
5) T	DICAMBA	12.187	11.424	-737369	5266316	N.D.	4.251
6) T	MCP	12.490	11.692	10229469	448150	9.033	0.467 #
7) T	MCPA	12.715	0.000	1437891	0	0.818	N.D. #
8) T	DICHLORPROP	13.300	12.621	1383093	544566	3.677	1.691 #
9) T	2,4-D	0.000	13.043	0	572745	N.D.	1.348 #
10) T	Pentachlo...	14.082	0.000	1711487	0	0.326	N.D. #
11) T	2,4,5-TP ...	0.000	14.261	0	850732	N.D.	0.437 #
12) T	2,4,5-T	0.000	14.851	0	5038532	N.D.	2.979 #
14) T	DINOSEB	17.572	15.920	8604183	613854	5.778	0.471 #
15) T	Picloram	17.349	17.171	5250211	4610648	1.798	1.848
16) T	DCPA	17.949	17.089	1875777	1222058	0.732	0.530 #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS062218\
Data File : PS000573.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Jun 2018 16:09
Operator : UA
Sample : J3603-18
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 22 17:47:13 2018
Quant Method : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061718.M
Quant Title : 8080.M
QLast Update : Mon Jun 18 00:27:44 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#2 3,5-DICHLOROBENZOIC ACID


R.T.: 0.000 min
Exp R.T.: 10.244 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 9.328 min
Delta R.T.: -0.020 min
Response: 763677
Conc: 1.66 ng/ml




#3 4-Nitrophenol

R.T.: 11.460 min
Delta R.T.: 0.033 min
Response: 4000969
Conc: 17.02 ng/ml



#3 4-Nitrophenol


R.T.: 10.338 min
Delta R.T.: -0.013 min
Response: 6519143
Conc: 33.27 ng/ml



#4 2,4-DCAA


R.T.: 11.802 min
Delta R.T.: -0.028 min
Response: 128224937
Conc: 340.43 ng/ml

Instrument :
ECD_S
ClientSampleId :




#4 2,4-DCAA

R.T.: 11.180 min
Delta R.T.: -0.023 min
Response: 105248211
Conc: 309.27 ng/ml




#5 DICAMBA

R.T.: 12.187 min
Delta R.T.: 0.040 min
Response: -737369
Conc: N.D.



#5 DICAMBA


R.T.: 11.424 min
Delta R.T.: -0.069 min
Response: 5266316
Conc: 4.25 ng/ml



#6 MCPP


R.T.: 12.490 min
Delta R.T.: 0.008 min
Response: 10229469
Conc: 9.03 ug/ml

Instrument :
ECD_S
ClientSampleId :




#6 MCPP

R.T.: 11.692 min
Delta R.T.: -0.019 min
Response: 448150
Conc: 0.47 ug/ml




#7 MCPA

R.T.: 12.715 min
Delta R.T.: -0.018 min
Response: 1437891
Conc: 0.82 ug/ml



#7 MCPA


R.T.: 0.000 min
Exp R.T. : 12.087 min
Response: 0
Conc: N.D.



#8 DICHLORPROP


R.T.: 13.300 min
Delta R.T.: -0.040 min
Response: 1383093
Conc: 3.68 ng/ml

Instrument :
ECD_S
ClientSampleId :




#8 DICHLORPROP

R.T.: 12.621 min
Delta R.T.: 0.005 min
Response: 544566
Conc: 1.69 ng/ml




#9 2,4-D

R.T.: 0.000 min
Exp R.T. : 13.699 min
Response: 0
Conc: N.D.



#9 2,4-D


R.T.: 13.043 min
Delta R.T.: -0.045 min
Response: 572745
Conc: 1.35 ng/ml



#10 Pentachlorophenol


R.T.: 14.082 min
Delta R.T.: -0.015 min
Response: 1711487
Conc: 0.33 ng/ml

Instrument :
ECD_S
ClientSampleId :




#10 Pentachlorophenol

R.T.: 0.000 min
Exp R.T. : 13.651 min
Response: 0
Conc: N.D.




#11 2,4,5-TP (SILVEX)

R.T.: 0.000 min
Exp R.T. : 14.962 min
Response: 0
Conc: N.D.



#11 2,4,5-TP (SILVEX)


R.T.: 14.261 min
Delta R.T.: 0.002 min
Response: 850732
Conc: 0.44 ng/ml



#12 2,4,5-T


R.T.: 0.000 min
Exp R.T.: 15.360 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#12 2,4,5-T

R.T.: 14.851 min
Delta R.T.: 0.016 min
Response: 5038532
Conc: 2.98 ng/ml




#14 DINOSEB

R.T.: 17.572 min
Delta R.T.: -0.054 min
Response: 8604183
Conc: 5.78 ng/ml



#14 DINOSEB


R.T.: 15.920 min
Delta R.T.: 0.009 min
Response: 613854
Conc: 0.47 ng/ml



#15 Picloram


R.T.: 17.349 min
Delta R.T.: -0.033 min
Response: 5250211
Conc: 1.80 ng/ml

Instrument :
ECD_S
ClientSampleId :




#15 Picloram

R.T.: 17.171 min
Delta R.T.: 0.015 min
Response: 4610648
Conc: 1.85 ng/ml



#16 DCPA

R.T.: 17.949 min
Delta R.T.: -0.016 min
Response: 1875777
Conc: 0.73 ng/ml



#16 DCPA

R.T.: 17.089 min
Delta R.T.: -0.009 min
Response: 1222058
Conc: 0.53 ng/ml