

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS062222\
 Data File : PS019767.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Jun 2022 00:28
 Operator : AJ\MA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 23 01:06:41 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052422.M
 Quant Title : 8080.M
 QLast Update : Wed May 25 02:05:34 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.194	6.779	1874.8E6	618.8E6	781.425	798.649
Target Compounds							
1) T	Dalapon	3.011	2.524	3130.5E6	690.5E6	617.609	688.163
2) T	3,5-DICHL...	6.409	5.960	2267.1E6	894.4E6	717.286	746.143
3) T	4-Nitroph...	7.025	6.397	738.3E6	304.2E6	752.368	710.715
5) T	DICAMBA	7.383	6.944	6967.8E6	2607.0E6	733.631	762.635
6) T	MCPD	7.554	7.042	551.0E6	204.8E6	73.409	75.658
7) T	MCPA	7.702	7.254	788.5E6	284.7E6	72.985	74.083
8) T	DICHLORPROP	8.074	7.563	1838.9E6	663.1E6	718.152	750.612
9) T	2,4-D	8.303	7.855	2123.2E6	795.8E6	742.524	746.096
10) T	Pentachlo...	8.610	8.280	25160.4E6	8794.6E6	727.412	754.497
11) T	2,4,5-TP ...	9.160	8.641	11448.3E6	3825.3E6	729.148	760.929
12) T	2,4,5-T	9.445	9.018	13377.4E6	3587.5E6	710.819	762.777
13) T	2,4-DB	9.997	9.523	3060.2E6	476.7E6	756.341	771.382
14) T	DINOSEB	11.150	9.848	11215.8E6	2501.9E6	725.137	704.252
15) T	Picloram	10.975	10.829	21853.1E6	4492.0E6	736.358	698.359
16) T	DCPA	11.419	10.776	18742.0E6	4700.6E6	777.547	761.543

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS062222\
Data File : PS019767.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jun 2022 00:28
Operator : AJ\MA
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 23 01:06:41 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052422.M
Quant Title : 8080.M
QLast Update : Wed May 25 02:05:34 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

