

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS062420\
 Data File : PS010905.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Jun 2020 20:31
 Operator : DD\AJ
 Sample : HSTDICC1500
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC1500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 25 02:34:40 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS062420.M
 Quant Title : 8080.M
 QLast Update : Thu Jun 25 02:30:34 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.228	6.920	1377.1E6	828.2E6	1514.652	1501.061
Target Compounds							
1) T	Dalapon	3.039	2.623	1562.1E6	1025.1E6	1385.834	1364.444
2) T	3,5-DICHL...	6.441	6.070	1752.0E6	1131.2E6	1385.355	1358.218
3) T	4-Nitroph...	7.059	6.525	559.2E6	432.2E6	1395.067	1345.470
5) T	DICAMBA	7.420	7.089	5037.3E6	2976.0E6	1365.227	1302.163
6) T	MCPD	7.596	7.194	430.2E6	277.2E6	148.631	146.189
7) T	MCPA	7.748	7.411	585.3E6	396.6E6	145.502	142.176
8) T	DICHLORPROP	8.113	7.715	1253.7E6	815.6E6	1424.516	1401.525
9) T	2,4-D	8.342	8.008	1295.2E6	1003.7E6	1445.962	1390.762
10) T	Pentachlo...	8.654	8.440	15762.4E6	7880.7E6	1260.848	1143.503
11) T	2,4,5-TP ...	9.203	8.800	6818.8E6	3892.2E6	1357.460	1284.938
12) T	2,4,5-T	9.488	9.179	6506.5E6	3910.2E6	1373.930	1297.546
13) T	2,4-DB	10.040	9.685	921.3E6	592.4E6	1528.160	1435.255
14) T	DINOSEB	11.202	10.019	4977.1E6	2807.8E6	1373.957	1331.485
15) T	Picloram	11.021	11.001	8588.5E6	5584.6E6	1464.310	1276.105
16) T	DCPA	11.468	10.949	8319.8E6	4747.5E6	1349.985	1271.655

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS062420\
Data File : PS010905.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Jun 2020 20:31
Operator : DD\AJ
Sample : HSTDICC1500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC1500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 25 02:34:40 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS062420.M
Quant Title : 8080.M
QLast Update : Thu Jun 25 02:30:34 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

