

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS062422\
 Data File : PS019802.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Jun 2022 07:50
 Operator : AJ\MA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 06/27/2022
 Supervised By :Ankita Jodhani 06/27/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 27 03:09:23 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052422.M
 Quant Title : 8080.M
 QLast Update : Wed May 25 02:05:34 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.193	6.779	2032.3E6	658.7E6	847.102	850.130
Target Compounds							
1) T	Dalapon	3.015	2.526	3534.4E6	733.6E6	697.291	731.109
2) T	3,5-DICHL...	6.408	5.959	2456.0E6	948.0E6	777.073	790.854
3) T	4-Nitroph...	7.024	6.396	892.7E6	334.8E6	909.708m	782.265
5) T	DICAMBA	7.382	6.943	7526.1E6	2752.0E6	792.403	805.046
6) T	MCPD	7.553	7.041	600.9E6	216.7E6	80.063	80.045
7) T	MCPA	7.701	7.254	859.7E6	301.2E6	79.576	78.395
8) T	DICHLORPROP	8.073	7.563	1989.8E6	701.1E6	777.072	793.703
9) T	2,4-D	8.303	7.854	2289.9E6	848.4E6	800.793	795.423
10) T	Pentachlo...	8.609	8.280	27036.1E6	9221.3E6	781.643	791.107
11) T	2,4,5-TP ...	9.159	8.640	11914.6E6	4029.4E6	758.848	801.528
12) T	2,4,5-T	9.444	9.017	13721.3E6	3809.0E6	729.094	809.874
13) T	2,4-DB	9.997	9.522	3209.0E6	506.7E6	793.114	819.953
14) T	DINOSEB	11.150	9.848	12119.6E6	2782.6E6	783.571	783.239
15) T	Picloram	10.975	10.829	22306.3E6	4558.3E6	751.629	708.662
16) T	DCPA	11.417	10.777	19470.9E6	4936.2E6	807.786	799.710

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS062422\
Data File : PS019802.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Jun 2022 07:50
Operator : AJ\MA
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 06/27/2022
Supervised By : Ankita Jodhani 06/27/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 27 03:09:23 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052422.M
Quant Title : 8080.M
QLast Update : Wed May 25 02:05:34 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

