

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS062520\
 Data File : PS010943.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Jun 2020 20:56
 Operator : DD\AJ
 Sample : L3089-08
 Misc :
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 RB47185-RT

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 26 01:41:33 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS062420.M
 Quant Title : 8080.M
 QLast Update : Thu Jun 25 02:45:12 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.229	6.920	372.6E6	253.8E6	410.887	458.072
Target Compounds							
1) T	Dalapon	2.997	2.583	36853634	74607828	32.913	98.902 #
2) T	3,5-DICHL...	0.000	6.100	0	70953380	N.D.	85.196
3) T	4-Nitroph...	6.997	6.551	2953541	16052648	7.399	49.956 #
5) T	DICAMBA	7.409	7.055	5142065	6591579	1.410	2.902 #
6) T	MCPD	7.583	0.000	6376833	0	2.236	N.D. #
7) T	MCPA	7.696	7.465	88740322	59551172	22.283	21.577
8) T	DICHLORPROP	8.061	7.740	14003349	23074167	15.969	39.300 #
10) T	Pentachlo...	8.709	0.000	13728670	0	1.104	N.D. #
12) T	2,4,5-T	9.428	0.000	19787495	0	4.248	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS062520\
Data File : PS010943.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Jun 2020 20:56
Operator : DD\AJ
Sample : L3089-08
Misc :
ALS Vial : 45 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
RB47185-RT

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 26 01:41:33 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS062420.M
Quant Title : 8080.M
QLast Update : Thu Jun 25 02:45:12 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 2.997 min
Delta R.T.: -0.041 min
Response: 36853634
Conc: 32.91 ng/ml

Instrument :
ECD_S
ClientSampleId :
RB47185-RT




#1 Dalapon

R.T.: 2.583 min
Delta R.T.: -0.038 min
Response: 74607828
Conc: 98.90 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 0.000 min
Exp R.T. : 6.441 min
Response: 0
Conc: N.D.



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 6.100 min
Delta R.T.: 0.030 min
Response: 70953380
Conc: 85.20 ng/ml



#3 4-Nitrophenol


R.T.: 6.997 min
Delta R.T.: -0.061 min
Response: 2953541
Conc: 7.40 ng/ml

Instrument :
ECD_S
ClientSampleId :
RB47185-RT




#3 4-Nitrophenol

R.T.: 6.551 min
Delta R.T.: 0.025 min
Response: 16052648
Conc: 49.96 ng/ml




#4 2,4-DCAA

R.T.: 7.229 min
Delta R.T.: 0.000 min
Response: 372599846
Conc: 410.89 ng/ml



#4 2,4-DCAA


R.T.: 6.920 min
Delta R.T.: 0.000 min
Response: 253807002
Conc: 458.07 ng/ml



#5 DICAMBA


R.T.: 7.409 min
Delta R.T.: -0.012 min
Response: 5142065
Conc: 1.41 ng/ml

Instrument :
ECD_S
ClientSampleId :
RB47185-RT




#5 DICAMBA

R.T.: 7.055 min
Delta R.T.: -0.035 min
Response: 6591579
Conc: 2.90 ng/ml




#6 MCP

R.T.: 7.583 min
Delta R.T.: -0.008 min
Response: 6376833
Conc: 2.24 ug/ml



#6 MCP


R.T.: 0.000 min
Exp R.T.: 7.187 min
Response: 0
Conc: N.D.



#7 MCPA


R.T.: 7.696 min
Delta R.T.: -0.044 min
Response: 88740322
Conc: 22.28 ug/ml

Instrument :
ECD_S
ClientSampleId :
RB47185-RT




#7 MCPA

R.T.: 7.465 min
Delta R.T.: 0.062 min
Response: 59551172
Conc: 21.58 ug/ml




#8 DICHLORPROP

R.T.: 8.061 min
Delta R.T.: -0.052 min
Response: 14003349
Conc: 15.97 ng/ml



#8 DICHLORPROP


R.T.: 7.740 min
Delta R.T.: 0.025 min
Response: 23074167
Conc: 39.30 ng/ml



#10 Pentachlorophenol


R.T.: 8.709 min
Delta R.T.: 0.055 min
Response: 13728670
Conc: 1.10 ng/ml

Instrument :
ECD_S
ClientSampleId :
RB47185-RT




#10 Pentachlorophenol

R.T.: 0.000 min
Exp R.T. : 8.440 min
Response: 0
Conc: N.D.



#12 2,4,5-T

R.T.: 9.428 min
Delta R.T.: -0.060 min
Response: 19787495
Conc: 4.25 ng/ml



#12 2,4,5-T

R.T.: 0.000 min
Exp R.T. : 9.179 min
Response: 0
Conc: N.D.