

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS062520\
 Data File : PS010912.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Jun 2020 09:05
 Operator : DD\AJ
 Sample : PB129715BSD
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PB129715BSD

Manual Integrations
 APPROVED

Ankita
 7/1/2020 11:31:58 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 25 13:44:49 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS062420.M
 Quant Title : 8080.M
 QLast Update : Thu Jun 25 02:45:12 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.229	6.921	411.4E6	263.6E6	453.655	475.729
Target Compounds							
1) T	Dalapon	3.033	2.620	98869904	116.1E6	88.299	153.929m#
2) T	3,5-DICHL...	6.442	6.070	582.6E6	373.5E6	463.125	448.458m
3) T	4-Nitroph...	7.060	6.526	172.2E6	130.7E6	431.355m	406.686m
5) T	DICAMBA	7.421	7.089	318.7E6	211.0E6	87.400	92.888m
6) T	MCPD	7.588	7.184	134.6E6	82667080	47.201	43.675
7) T	MCPA	7.737	7.398	173.4E6	149.8E6	43.539	54.273
8) T	DICHLORPROP	8.113	7.716	385.7E6	258.4E6	439.779	440.081
9) T	2,4-D	8.342	8.010	424.9E6	350.4E6	480.467	484.841
10) T	Pentachlo...	8.653	8.440	6169.1E6	3521.6E6	496.138	506.436
11) T	2,4,5-TP ...	9.202	8.800	2410.1E6	1467.8E6	485.380	487.820
12) T	2,4,5-T	9.488	9.179	2105.5E6	1402.7E6	452.031	471.035
13) T	2,4-DB	10.041	9.686	241.0E6	188.7E6	405.562	459.131
14) T	DINOSEB	11.203	10.019	1565.0E6	966.8E6	436.961	464.336
15) T	Picloram	11.021	11.006	339.6E6	145.6E6	60.573	34.741 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS062520\
Data File : PS010912.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Jun 2020 09:05
Operator : DD\AJ
Sample : PB129715BSD
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
PB129715BSD

Manual Integrations
APPROVED

Ankita
7/1/2020 11:31:58 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 25 13:44:49 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS062420.M
Quant Title : 8080.M
QLast Update : Thu Jun 25 02:45:12 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

