

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS062523\
 Data File : PS023398.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Jun 2023 12:34
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 26 01:16:46 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS060623.M
 Quant Title : 8080.M
 QLast Update : Tue Jun 06 13:00:04 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.479	6.852	1379.2E6	444.4E6	779.511	782.784
Target Compounds							
1) T	Dalapon	2.549	2.575	1275.7E6	601.6E6	697.654	656.281
2) T	3,5-DICHL...	5.833	6.020	1961.7E6	654.3E6	730.199	739.180
3) T	4-Nitroph...	6.312	6.458	935.9E6	239.2E6	726.141	701.031
5) T	DICAMBA	6.628	7.019	5851.2E6	1938.5E6	750.759	756.442
6) T	MCPD	6.786	7.114	426.9E6	134.7E6	77.448	75.369
7) T	MCPA	6.914	7.325	581.4E6	193.5E6	74.259	71.982
8) T	DICHLORPROP	7.236	7.641	1499.4E6	457.7E6	726.609	739.960
9) T	2,4-D	7.438	7.928	1776.0E6	514.0E6	726.903	741.117
10) T	Pentachlo...	7.672	8.359	21055.6E6	6872.5E6	747.199	759.653
11) T	2,4,5-TP ...	8.208	8.718	8688.5E6	2547.6E6	751.391	763.359
12) T	2,4,5-T	8.468	9.091	8467.1E6	2316.4E6	745.326	755.631
13) T	2,4-DB	8.983	9.597	1684.9E6	285.7E6	715.726	770.358
14) T	DINOSEB	10.056	9.927	6079.5E6	1782.5E6	734.048	751.131
15) T	Picloram	9.877	10.890	12466.3E6	3266.5E6	718.332	715.747
16) T	DCPA	10.326	10.854	10467.5E6	3075.0E6	750.867	766.468

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS062523\
Data File : PS023398.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Jun 2023 12:34
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 26 01:16:46 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS060623.M
Quant Title : 8080.M
QLast Update : Tue Jun 06 13:00:04 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

