

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS062523\
 Data File : PS023420.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Jun 2023 22:56
 Operator : AR\AJ
 Sample : 03330-07MSD
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 COMPOSITE-MC01MSD

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 06/26/2023
 Supervised By :Ankita Jodhani 06/27/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 26 01:25:36 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS060623.M
 Quant Title : 8080.M
 QLast Update : Tue Jun 06 13:00:04 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.477	6.852	340.3E6	89919673	192.329m	158.394
Target Compounds							
1) T	Dalapon	2.547	2.574	231.9E6	200.4E6	126.814	218.668 #
2) T	3,5-DICHL...	5.832	6.019	365.5E6	112.0E6	136.049	126.574
5) T	DICAMBA	6.626	7.018	1130.7E6	320.9E6	145.072	125.222
6) T	MCP	6.781	7.110	50785015	24519765	9.214	13.716 #
7) T	MCPA	6.908	7.320	103.2E6	45913402	13.175m	17.083 #
8) T	DICHLORPROP	7.234	7.640	253.2E6	90287243	122.691m	145.957
9) T	2,4-D	7.436	7.927	423.2E6	120.6E6	173.225	173.882
10) T	Pentachlo...	7.670	8.358	2849.6E6	836.0E6	101.122	92.408
11) T	2,4,5-TP ...	8.205	8.717	1252.3E6	345.7E6	108.300	103.591
12) T	2,4,5-T	8.465	9.091	1503.7E6	376.9E6	132.365	122.961
13) T	2,4-DB	8.983	9.597	374.9E6	53364660	159.271	143.917
15) T	Picloram	9.874	10.889	2886.7E6	710.5E6	166.335	155.679
16) T	DCPA	10.322	10.853	2393.6E6	633.8E6	171.701	157.983

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS062523\
Data File : PS023420.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Jun 2023 22:56
Operator : AR\AJ
Sample : 03330-07MSD
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
COMPOSITE-MC01MSD

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 06/26/2023
Supervised By : Ankita Jodhani 06/27/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 26 01:25:36 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS060623.M
Quant Title : 8080.M
QLast Update : Tue Jun 06 13:00:04 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

