

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS062523\
 Data File : PS023428.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Jun 2023 03:23
 Operator : AR\AJ
 Sample : 03357-02MS
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 WC-SCAN-01MS

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 06/26/2023
 Supervised By : Ankita Jodhani 06/27/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 26 07:17:04 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS060623.M
 Quant Title : 8080.M
 QLast Update : Tue Jun 06 13:00:04 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.481	6.852	1183.0E6	286.4E6	668.640	504.488
Target Compounds							
1) T	Dalapon	2.549	2.576	632.9E6	401.9E6	346.142	438.488 #
2) T	3,5-DICHL...	5.832	6.019	1314.9E6	417.3E6	489.446	471.373
5) T	DICAMBA	6.626	7.018	3498.3E6	1140.3E6	448.858	444.941
6) T	MCP	6.784	7.113	287.2E6	98937035	52.112	55.344
7) T	MCPA	6.911	7.322	335.1E6	145.2E6	42.803	54.014 #
8) T	DICHLORPROP	7.235	7.640	1010.7E6	297.8E6	489.777	481.456
9) T	2,4-D	7.436	7.927	1174.4E6	343.1E6	480.667	494.719
10) T	Pentachlo...	7.671	8.357	13898.9E6	4393.2E6	493.230	485.604
11) T	2,4,5-TP ...	8.206	8.717	5693.0E6	1730.1E6	492.336	518.403
12) T	2,4,5-T	8.466	9.090	5581.9E6	1522.8E6	491.360	496.735
13) T	2,4-DB	8.983	9.597	891.8E6	200.4E6	378.836	540.403 #
14) T	DINOSEB	10.054	9.925	3390.9E6	1127.4E6	409.422m	475.068
15) T	Picloram	9.874	10.888	7168.2E6	1835.6E6	413.045	402.208
16) T	DCPA	10.323	10.852	6773.9E6	1973.8E6	485.915	491.982

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS062523\
Data File : PS023428.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Jun 2023 03:23
Operator : AR\AJ
Sample : 03357-02MS
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
WC-SCAN-01MS

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 06/26/2023
Supervised By : Ankita Jodhani 06/27/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 26 07:17:04 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS060623.M
Quant Title : 8080.M
QLast Update : Tue Jun 06 13:00:04 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

