

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS062618\
 Data File : PS000647.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Jun 2018 17:40
 Operator : UA
 Sample : J3573-09
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 27 00:48:22 2018
 Quant Method : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061718.M
 Quant Title : 8080.M
 QLast Update : Mon Jun 18 00:27:44 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.863	11.175	1719851	109.3E6	4.566	321.150 #
Target Compounds							
1) T	Dalapon	0.000	2.598	0	8306069	N.D.	19.054 #
2) T	3,5-DICHL...	10.307	9.323	11054	2783487	0.020	6.051 #
3) T	4-Nitroph...	11.455	10.333	1816135	9900797	7.726	50.524 #
5) T	DICAMBA	12.185	11.500	3291359	789021	2.329	0.637 #
6) T	MCPD	12.485	11.685	9622154	700466	8.497	0.730 #
7) T	MCPA	12.707	12.095	975489	36330000	0.555	26.232 #
8) T	DICHLORPROP	13.341	12.616	1242144	2533210	3.302	7.868 #
9) T	2,4-D	0.000	13.135	0	555765	N.D.	1.308 #
10) T	Pentachlo...	14.062	13.626	6186511	3267133	1.178	0.691 #
11) T	2,4,5-TP ...	0.000	14.255	0	4563042	N.D.	2.345 #
12) T	2,4,5-T	0.000	14.849	0	5300475	N.D.	3.134 #
13) T	2,4-DB	16.120	15.459	434821	2800708	1.477	11.314 #
14) T	DINOSEB	17.628	15.947	2627037	18623925	1.764	14.276 #
15) T	Picloram	17.412	17.170	1794642	7257326	0.615	2.909 #
16) T	DCPA	17.937	17.076	3380816	5976077	1.319	2.592 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS062618\
Data File : PS000647.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Jun 2018 17:40
Operator : UA
Sample : J3573-09
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 27 00:48:22 2018
Quant Method : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061718.M
Quant Title : 8080.M
QLast Update : Mon Jun 18 00:27:44 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 0.000 min
Exp R.T.: 3.182 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#1 Dalapon

R.T.: 2.598 min
Delta R.T.: 0.038 min
Response: 8306069
Conc: 19.05 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 10.307 min
Delta R.T.: 0.063 min
Response: 11054
Conc: 0.02 ng/ml



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 9.323 min
Delta R.T.: -0.024 min
Response: 2783487
Conc: 6.05 ng/ml



#3 4-Nitrophenol


R.T.: 11.455 min
Delta R.T.: 0.027 min
Response: 1816135
Conc: 7.73 ng/ml

Instrument :
ECD_S
ClientSampleId :




#3 4-Nitrophenol

R.T.: 10.333 min
Delta R.T.: -0.018 min
Response: 9900797
Conc: 50.52 ng/ml




#4 2,4-DCAA

R.T.: 11.863 min
Delta R.T.: 0.034 min
Response: 1719851
Conc: 4.57 ng/ml



#4 2,4-DCAA


R.T.: 11.175 min
Delta R.T.: -0.028 min
Response: 109290106
Conc: 321.15 ng/ml



#5 DICAMBA


R.T.: 12.185 min
Delta R.T.: 0.037 min
Response: 3291359
Conc: 2.33 ng/ml

Instrument :
ECD_S
ClientSampleId :




#5 DICAMBA

R.T.: 11.500 min
Delta R.T.: 0.008 min
Response: 789021
Conc: 0.64 ng/ml




#6 MCP

R.T.: 12.485 min
Delta R.T.: 0.003 min
Response: 9622154
Conc: 8.50 ug/ml



#6 MCP


R.T.: 11.685 min
Delta R.T.: -0.027 min
Response: 700466
Conc: 0.73 ug/ml



#7 MCPA


R.T.: 12.707 min
Delta R.T.: -0.026 min
Response: 975489
Conc: 0.55 ug/ml

Instrument :
ECD_S
ClientSampleId :




#7 MCPA

R.T.: 12.095 min
Delta R.T.: 0.008 min
Response: 36330000
Conc: 26.23 ug/ml




#8 DICHLORPROP

R.T.: 13.341 min
Delta R.T.: 0.000 min
Response: 1242144
Conc: 3.30 ng/ml



#8 DICHLORPROP


R.T.: 12.616 min
Delta R.T.: 0.000 min
Response: 2533210
Conc: 7.87 ng/ml



#9 2,4-D


R.T.: 0.000 min
Exp R.T.: 13.699 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#9 2,4-D

R.T.: 13.135 min
Delta R.T.: 0.048 min
Response: 555765
Conc: 1.31 ng/ml




#10 Pentachlorophenol

R.T.: 14.062 min
Delta R.T.: -0.034 min
Response: 6186511
Conc: 1.18 ng/ml



#10 Pentachlorophenol


R.T.: 13.626 min
Delta R.T.: -0.024 min
Response: 3267133
Conc: 0.69 ng/ml



#11 2,4,5-TP (SILVEX)


R.T.: 0.000 min
Exp R.T.: 14.962 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#11 2,4,5-TP (SILVEX)

R.T.: 14.255 min
Delta R.T.: -0.004 min
Response: 4563042
Conc: 2.34 ng/ml




#12 2,4,5-T

R.T.: 0.000 min
Exp R.T.: 15.360 min
Response: 0
Conc: N.D.



#12 2,4,5-T


R.T.: 14.849 min
Delta R.T.: 0.015 min
Response: 5300475
Conc: 3.13 ng/ml



#13 2,4-DB


R.T.: 16.120 min
Delta R.T.: 0.006 min
Response: 434821
Conc: 1.48 ng/ml

Instrument :
ECD_S
ClientSampleId :




#13 2,4-DB

R.T.: 15.459 min
Delta R.T.: -0.019 min
Response: 2800708
Conc: 11.31 ng/ml




#14 DINOSEB

R.T.: 17.628 min
Delta R.T.: 0.002 min
Response: 2627037
Conc: 1.76 ng/ml



#14 DINOSEB


R.T.: 15.947 min
Delta R.T.: 0.036 min
Response: 18623925
Conc: 14.28 ng/ml



#15 Picloram


R.T.: 17.412 min
Delta R.T.: 0.030 min
Response: 1794642
Conc: 0.61 ng/ml

Instrument :
ECD_S
ClientSampleId :




#15 Picloram

R.T.: 17.170 min
Delta R.T.: 0.014 min
Response: 7257326
Conc: 2.91 ng/ml



#16 DCPA

R.T.: 17.937 min
Delta R.T.: -0.027 min
Response: 3380816
Conc: 1.32 ng/ml



#16 DCPA

R.T.: 17.076 min
Delta R.T.: -0.022 min
Response: 5976077
Conc: 2.59 ng/ml