

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS062619\
 Data File : PS005239.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Jun 2019 18:10
 Operator : SM\AJ
 Sample : PB120891BS
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PB120891BS

Manual Integrations
 APPROVED

Ankita
 6/27/2019 9:12:14 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 27 00:41:29 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061919.M
 Quant Title : 8080.M
 QLast Update : Thu Jun 20 11:04:04 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.588	5.302	28753736	20716530	19.588m	20.668m
Target Compounds							
1) T	Dalapon	2.093	0.000	46215159	0	25.072	N.D. #
2) T	3,5-DICHL...	4.904	4.559	2574996	8492566	1.281m	6.440m#
3) T	4-Nitroph...	5.389	4.964	13946047	5030538	19.255	10.032m#
5) T	DICAMBA	5.746	5.434	30009343	25091545	5.260m	6.703m#
7) T	MCPA	6.079	5.692	26869478	26775110	4.279m	6.610 #
8) T	DICHLORPROP	6.386	5.991	10783196	12879948	7.540m	13.340m#
9) T	2,4-D	6.621	6.249	10340491	9134102	6.652m	8.190m
10) T	Pentachlo...	6.872	6.639	83858877	61111133	4.262m	4.753
11) T	2,4,5-TP ...	7.410	7.001	35474705	26154300	4.348	4.911
12) T	2,4,5-T	7.687	7.367	46782481	21408356	5.813	4.197m#
13) T	2,4-DB	8.230	7.861	285.5E6	232.1E6	245.773	355.588 #
14) T	DINOSEB	9.314	8.161	7761567	43007114	1.214m	10.637 #
15) T	Picloram	9.162	9.118	24229380	17817764	2.109m	2.361
16) T	DCPA	9.584	9.066	50722496	14648017	5.085m	2.152m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS062619\
Data File : PS005239.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Jun 2019 18:10
Operator : SM\AJ
Sample : PB120891BS
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
PB120891BS

Manual Integrations
APPROVED

Ankita
6/27/2019 9:12:14 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 27 00:41:29 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061919.M
Quant Title : 8080.M
QLast Update : Thu Jun 20 11:04:04 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

