

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS062620\
 Data File : PS010961.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Jun 2020 16:16
 Operator : DD\AJ
 Sample : PB129795BL
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PB129795BL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 27 03:57:42 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS062420.M
 Quant Title : 8080.M
 QLast Update : Thu Jun 25 02:45:12 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.229	6.920	442.5E6	304.0E6	487.970	548.572
Target Compounds							
1) T	Dalapon	2.994	2.584	42760323	182.9E6	38.188	242.446 #
2) T	3,5-DICHL...	0.000	6.100	0	31133412	N.D.	37.383
3) T	4-Nitroph...	7.006	0.000	692268	0	1.734	N.D. #
7) T	MCPA	7.777	7.384	46004387	36104461	11.552	13.081
8) T	DICHLORPROP	0.000	7.739	0	34820943	N.D.	59.307
10) T	Pentachlo...	8.709	0.000	14636097	0	1.177	N.D. #
12) T	2,4,5-T	9.510	9.237	7159289	10080792	1.537	3.385 #
14) T	DINOSEB	11.168	10.038	58067611	41218969	16.213	19.796
15) T	Picloram	10.994	11.005	859.2E6	100.5E6	153.264	23.996 #
16) T	DCPA	0.000	11.005	0	100.5E6	N.D.	26.993



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS062620\
Data File : PS010961.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Jun 2020 16:16
Operator : DD\AJ
Sample : PB129795BL
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
PB129795BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 27 03:57:42 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS062420.M
Quant Title : 8080.M
QLast Update : Thu Jun 25 02:45:12 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 2.994 min
Delta R.T.: -0.044 min
Response: 42760323
Conc: 38.19 ng/ml

Instrument :
ECD_S
ClientSampleId :
PB129795BL




#1 Dalapon

R.T.: 2.584 min
Delta R.T.: -0.038 min
Response: 182891998
Conc: 242.45 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 0.000 min
Exp R.T. : 6.441 min
Response: 0
Conc: N.D.



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 6.100 min
Delta R.T.: 0.030 min
Response: 31133412
Conc: 37.38 ng/ml



#3 4-Nitrophenol


R.T.: 7.006 min
Delta R.T.: -0.053 min
Response: 692268
Conc: 1.73 ng/ml

Instrument :
ECD_S
ClientSampleId :
PB129795BL




#3 4-Nitrophenol

R.T.: 0.000 min
Exp R.T.: 6.526 min
Response: 0
Conc: N.D.




#4 2,4-DCAA

R.T.: 7.229 min
Delta R.T.: 0.000 min
Response: 442500364
Conc: 487.97 ng/ml



#4 2,4-DCAA


R.T.: 6.920 min
Delta R.T.: 0.000 min
Response: 303950701
Conc: 548.57 ng/ml



#7 MCPA


R.T.: 7.777 min
Delta R.T.: 0.037 min
Response: 46004387
Conc: 11.55 ug/ml

Instrument :
ECD_S
ClientSampleId :
PB129795BL




#7 MCPA

R.T.: 7.384 min
Delta R.T.: -0.019 min
Response: 36104461
Conc: 13.08 ug/ml




#8 DICHLORPROP

R.T.: 0.000 min
Exp R.T. : 8.113 min
Response: 0
Conc: N.D.



#8 DICHLORPROP


R.T.: 7.739 min
Delta R.T.: 0.024 min
Response: 34820943
Conc: 59.31 ng/ml



#10 Pentachlorophenol


R.T.: 8.709 min
Delta R.T.: 0.056 min
Response: 14636097
Conc: 1.18 ng/ml

Instrument :
ECD_S
ClientSampleId :
PB129795BL




#10 Pentachlorophenol

R.T.: 0.000 min
Exp R.T. : 8.440 min
Response: 0
Conc: N.D.




#11 2,4,5-TP (SILVEX)

R.T.: 9.267 min
Delta R.T.: 0.065 min
Response: 25687626
Conc: 5.17 ng/ml



#11 2,4,5-TP (SILVEX)


R.T.: 8.843 min
Delta R.T.: 0.043 min
Response: -372320276
Conc: N.D.



#12 2,4,5-T


R.T.: 9.510 min
Delta R.T.: 0.022 min
Response: 7159289
Conc: 1.54 ng/ml

Instrument :
ECD_S
ClientSampleId :
PB129795BL




#12 2,4,5-T

R.T.: 9.237 min
Delta R.T.: 0.058 min
Response: 10080792
Conc: 3.39 ng/ml




#14 DINOSEB

R.T.: 11.168 min
Delta R.T.: -0.034 min
Response: 58067611
Conc: 16.21 ng/ml



#14 DINOSEB


R.T.: 10.038 min
Delta R.T.: 0.019 min
Response: 41218969
Conc: 19.80 ng/ml



#15 Picloram


R.T.: 10.994 min
Delta R.T.: -0.029 min
Response: 859218790
Conc: 153.26 ng/ml

Instrument :
ECD_S
ClientSampleId :
PB129795BL




#15 Picloram

R.T.: 11.005 min
Delta R.T.: 0.003 min
Response: 100534352
Conc: 24.00 ng/ml



#16 DCPA

R.T.: 0.000 min
Exp R.T. : 11.468 min
Response: 0
Conc: N.D.



#16 DCPA

R.T.: 11.005 min
Delta R.T.: 0.056 min
Response: 100534352
Conc: 26.99 ng/ml