

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS062620\
 Data File : PS010986.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Jun 2020 01:50
 Operator : DD\AJ
 Sample : L3114-02
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 VNJ-242

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 27 04:06:22 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS062420.M
 Quant Title : 8080.M
 QLast Update : Thu Jun 25 02:45:12 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.229	6.919	248.7E6	172.4E6	274.257	311.212
Target Compounds							
1) T	Dalapon	2.995	2.586	38351489	133.8E6	34.251	177.316 #
2) T	3,5-DICHL...	0.000	6.100	0	64725457	N.D.	77.718
3) T	4-Nitroph...	7.118	0.000	9240804	0	23.150	N.D. #
5) T	DICAMBA	0.000	7.083	0	11127812	N.D.	4.899
7) T	MCPA	7.683	7.453	258.6E6	205.2E6	64.940	74.352
8) T	DICHLORPROP	0.000	7.738	0	17040549	N.D.	29.023
12) T	2,4,5-T	9.433	0.000	59914162	0	12.863	N.D. #
14) T	DINOSEB	11.166	10.061	48865372	270.7E6	13.644	130.008 #
15) T	Picloram	10.991	0.000	866.8E6	0	154.613	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS062620\
Data File : PS010986.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Jun 2020 01:50
Operator : DD\AJ
Sample : L3114-02
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
VNJ-242

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 27 04:06:22 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS062420.M
Quant Title : 8080.M
QLast Update : Thu Jun 25 02:45:12 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 2.995 min
Delta R.T.: -0.043 min
Response: 38351489
Conc: 34.25 ng/ml

Instrument :
ECD_S
ClientSampleId :
VNJ-242




#1 Dalapon

R.T.: 2.586 min
Delta R.T.: -0.036 min
Response: 133760610
Conc: 177.32 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 0.000 min
Exp R.T. : 6.441 min
Response: 0
Conc: N.D.



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 6.100 min
Delta R.T.: 0.030 min
Response: 64725457
Conc: 77.72 ng/ml



#3 4-Nitrophenol


R.T.: 7.118 min
Delta R.T.: 0.059 min
Response: 9240804
Conc: 23.15 ng/ml

Instrument :
ECD_S
ClientSampleId :
VNJ-242




#3 4-Nitrophenol

R.T.: 0.000 min
Exp R.T.: 6.526 min
Response: 0
Conc: N.D.




#4 2,4-DCAA

R.T.: 7.229 min
Delta R.T.: 0.000 min
Response: 248701738
Conc: 274.26 ng/ml



#4 2,4-DCAA


R.T.: 6.919 min
Delta R.T.: 0.000 min
Response: 172435368
Conc: 311.21 ng/ml



#5 DICAMBA


R.T.: 0.000 min
Exp R.T.: 7.420 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :
VNJ-242




#5 DICAMBA

R.T.: 7.083 min
Delta R.T.: -0.006 min
Response: 11127812
Conc: 4.90 ng/ml




#7 MCPA

R.T.: 7.683 min
Delta R.T.: -0.057 min
Response: 258619781
Conc: 64.94 ug/ml



#7 MCPA


R.T.: 7.453 min
Delta R.T.: 0.050 min
Response: 205211491
Conc: 74.35 ug/ml



#8 DICHLORPROP


R.T.: 0.000 min
Exp R.T.: 8.113 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :
VNJ-242




#8 DICHLORPROP

R.T.: 7.738 min
Delta R.T.: 0.022 min
Response: 17040549
Conc: 29.02 ng/ml




#11 2,4,5-TP (SILVEX)

R.T.: 9.239 min
Delta R.T.: 0.036 min
Response: 9201622
Conc: 1.85 ng/ml



#11 2,4,5-TP (SILVEX)


R.T.: 8.841 min
Delta R.T.: 0.041 min
Response: -76961398
Conc: N.D.



#12 2,4,5-T


R.T.: 9.433 min
Delta R.T.: -0.055 min
Response: 59914162
Conc: 12.86 ng/ml

Instrument :
ECD_S
ClientSampleId :
VNJ-242




#12 2,4,5-T

R.T.: 0.000 min
Exp R.T. : 9.179 min
Response: 0
Conc: N.D.




#14 DINOSEB

R.T.: 11.166 min
Delta R.T.: -0.036 min
Response: 48865372
Conc: 13.64 ng/ml



#14 DINOSEB


R.T.: 10.061 min
Delta R.T.: 0.041 min
Response: 270694283
Conc: 130.01 ng/ml



#15 Picloram

R.T.: 10.991 min
Delta R.T.: -0.031 min
Response: 866780475
Conc: 154.61 ng/ml

Instrument :
ECD_S
ClientSampleId :
VNJ-242



#15 Picloram

R.T.: 0.000 min
Exp R.T. : 11.002 min
Response: 0
Conc: N.D.