

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS062620\
 Data File : PS010970.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Jun 2020 19:43
 Operator : DD\AJ
 Sample : PB129801BS
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PB129801BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 27 04:01:09 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS062420.M
 Quant Title : 8080.M
 QLast Update : Thu Jun 25 02:45:12 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.228	6.920	452.4E6	280.1E6	498.910	505.504
Target Compounds							
1) T	Dalapon	3.041	2.625	494.5E6	327.5E6	441.623	434.105
2) T	3,5-DICHL...	6.441	6.070	581.4E6	395.5E6	462.124	474.891
3) T	4-Nitroph...	7.058	6.525	197.7E6	153.0E6	495.163	476.030
5) T	DICAMBA	7.420	7.089	1741.4E6	1106.3E6	477.567	487.094
6) T	MCPD	7.588	7.184	131.4E6	88821086	46.093	46.926
7) T	MCPA	7.737	7.399	186.9E6	128.0E6	46.931	46.372
8) T	DICHLORPROP	8.111	7.715	408.6E6	279.1E6	465.907	475.335
9) T	2,4-D	8.341	8.009	429.0E6	338.5E6	485.102	468.424
10) T	Pentachlo...	8.653	8.440	6169.6E6	3644.2E6	496.182	524.069
11) T	2,4,5-TP ...	9.202	8.800	2384.2E6	1496.9E6	480.167	497.497
12) T	2,4,5-T	9.487	9.179	2278.1E6	1495.1E6	489.090	502.085
13) T	2,4-DB	10.039	9.685	291.1E6	197.5E6	489.870	480.666
14) T	DINOSEB	11.202	10.019	1642.7E6	1017.7E6	458.673	488.756
15) T	Picloram	11.020	11.001	2049.2E6	1469.7E6	365.535	350.811
16) T	DCPA	11.466	10.948	2930.2E6	1880.7E6	479.303	504.969

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS062620\
Data File : PS010970.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Jun 2020 19:43
Operator : DD\AJ
Sample : PB129801BS
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
PB129801BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 27 04:01:09 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS062420.M
Quant Title : 8080.M
QLast Update : Thu Jun 25 02:45:12 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

