

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS062718\
 Data File : PS000657.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Jun 2018 09:14
 Operator : UA
 Sample : HEXANE
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 27 15:04:26 2018
 Quant Method : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061718.M
 Quant Title : 8080.M
 QLast Update : Mon Jun 18 00:27:44 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.803	11.173	991423	641114	2.632	1.884 #
Target Compounds							
1) T	Dalapon	0.000	2.562	0	64210	N.D.	0.147 #
2) T	3,5-DICHL...	10.216	9.314	357138	31483	0.662	0.068 #
3) T	4-Nitroph...	11.402	10.318	604364	334503	2.571	1.707 #
5) T	DICAMBA	12.122	11.461	2409412	1956983	1.705	1.580
6) T	MCP	12.493	11.718	-1126341	964531	N.D.	1.005
8) T	DICHLORPROP	13.314	12.585	1041760	958662	2.769	2.977
9) T	2,4-D	13.673	13.059	1648779	1442941	3.447	3.395
10) T	Pentachlo...	14.069	13.621	10973646	8550370	2.090	1.807
11) T	2,4,5-TP ...	14.935	14.230	6899938	5674577	3.178	2.916
12) T	2,4,5-T	15.333	14.806	7912591	5391652	3.675	3.188
13) T	2,4-DB	16.086	15.449	794159	784619	2.697	3.170
14) T	DINOSEB	17.597	15.884	6371951	4711433	4.279	3.611
15) T	Picloram	17.355	17.129	10503353	7460080	3.596	2.991
16) T	DCPA	17.935	17.070	9410781	7638961	3.673	3.313



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS062718\
Data File : PS000657.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Jun 2018 09:14
Operator : UA
Sample : HEXANE
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 27 15:04:26 2018
Quant Method : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061718.M
Quant Title : 8080.M
QLast Update : Mon Jun 18 00:27:44 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 0.000 min
Exp R.T.: 3.182 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#1 Dalapon

R.T.: 2.562 min
Delta R.T.: 0.001 min
Response: 64210
Conc: 0.15 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 10.216 min
Delta R.T.: -0.027 min
Response: 357138
Conc: 0.66 ng/ml



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 9.314 min
Delta R.T.: -0.033 min
Response: 31483
Conc: 0.07 ng/ml



#3 4-Nitrophenol


R.T.: 11.402 min
Delta R.T.: -0.026 min
Response: 604364
Conc: 2.57 ng/ml

Instrument :
ECD_S
ClientSampleId :




#3 4-Nitrophenol

R.T.: 10.318 min
Delta R.T.: -0.032 min
Response: 334503
Conc: 1.71 ng/ml




#4 2,4-DCAA

R.T.: 11.803 min
Delta R.T.: -0.026 min
Response: 991423
Conc: 2.63 ng/ml



#4 2,4-DCAA


R.T.: 11.173 min
Delta R.T.: -0.030 min
Response: 641114
Conc: 1.88 ng/ml



#5 DICAMBA


R.T.: 12.122 min
Delta R.T.: -0.026 min
Response: 2409412
Conc: 1.70 ng/ml

Instrument :
ECD_S
ClientSampleId :




#5 DICAMBA

R.T.: 11.461 min
Delta R.T.: -0.032 min
Response: 1956983
Conc: 1.58 ng/ml




#6 MCP

R.T.: 12.493 min
Delta R.T.: 0.011 min
Response: -1126341
Conc: N.D.



#6 MCP


R.T.: 11.718 min
Delta R.T.: 0.007 min
Response: 964531
Conc: 1.01 ug/ml



#8 DICHLORPROP


R.T.: 13.314 min
Delta R.T.: -0.027 min
Response: 1041760
Conc: 2.77 ng/ml

Instrument :
ECD_S
ClientSampleId :




#8 DICHLORPROP

R.T.: 12.585 min
Delta R.T.: -0.031 min
Response: 958662
Conc: 2.98 ng/ml




#9 2,4-D

R.T.: 13.673 min
Delta R.T.: -0.026 min
Response: 1648779
Conc: 3.45 ng/ml



#9 2,4-D


R.T.: 13.059 min
Delta R.T.: -0.028 min
Response: 1442941
Conc: 3.40 ng/ml



#10 Pentachlorophenol


R.T.: 14.069 min
Delta R.T.: -0.028 min
Response: 10973646
Conc: 2.09 ng/ml

Instrument :
ECD_S
ClientSampleId :




#10 Pentachlorophenol

R.T.: 13.621 min
Delta R.T.: -0.030 min
Response: 8550370
Conc: 1.81 ng/ml




#11 2,4,5-TP (SILVEX)

R.T.: 14.935 min
Delta R.T.: -0.027 min
Response: 6899938
Conc: 3.18 ng/ml



#11 2,4,5-TP (SILVEX)


R.T.: 14.230 min
Delta R.T.: -0.029 min
Response: 5674577
Conc: 2.92 ng/ml



#12 2,4,5-T


R.T.: 15.333 min
Delta R.T.: -0.027 min
Response: 7912591
Conc: 3.67 ng/ml

Instrument :
ECD_S
ClientSampleId :




#12 2,4,5-T

R.T.: 14.806 min
Delta R.T.: -0.029 min
Response: 5391652
Conc: 3.19 ng/ml




#13 2,4-DB

R.T.: 16.086 min
Delta R.T.: -0.028 min
Response: 794159
Conc: 2.70 ng/ml



#13 2,4-DB


R.T.: 15.449 min
Delta R.T.: -0.029 min
Response: 784619
Conc: 3.17 ng/ml



#14 DINOSEB


R.T.: 17.597 min
Delta R.T.: -0.029 min
Response: 6371951
Conc: 4.28 ng/ml

Instrument :
ECD_S
ClientSampleId :




#14 DINOSEB

R.T.: 15.884 min
Delta R.T.: -0.028 min
Response: 4711433
Conc: 3.61 ng/ml




#15 Picloram

R.T.: 17.355 min
Delta R.T.: -0.027 min
Response: 10503353
Conc: 3.60 ng/ml



#15 Picloram


R.T.: 17.129 min
Delta R.T.: -0.027 min
Response: 7460080
Conc: 2.99 ng/ml



#16 DCPA

R.T.: 17.935 min
Delta R.T.: -0.029 min
Response: 9410781
Conc: 3.67 ng/ml

Instrument :
ECD_S
ClientSampleId :



#16 DCPA

R.T.: 17.070 min
Delta R.T.: -0.028 min
Response: 7638961
Conc: 3.31 ng/ml