

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS062818\
 Data File : PS000698.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Jun 2018 11:14
 Operator : UA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 29 00:55:06 2018
 Quant Method : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061718.M
 Quant Title : 8080.M
 QLast Update : Mon Jun 18 00:27:44 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.814	11.195	285.4E6	261.6E6	757.691	768.731
Target Compounds							
1) T	Dalapon	3.174	2.554	349.5E6	301.3E6	683.258	691.093
2) T	3,5-DICHL...	10.228	9.340	386.4E6	331.5E6	716.014	720.580
3) T	4-Nitroph...	11.412	10.342	161.3E6	135.5E6	686.204	691.399
5) T	DICAMBA	12.133	11.484	1047.1E6	931.1E6	740.803	751.479
6) T	MCPD	12.472	11.709	81736569	68568702	72.179	71.466
7) T	MCPA	12.724	12.088	122.9E6	103.5E6	69.942	74.752
8) T	DICHLORPROP	13.325	12.608	269.7E6	234.1E6	716.877	727.172
9) T	2,4-D	13.683	13.079	349.4E6	316.2E6	730.509	743.948
10) T	Pentachlo...	14.081	13.642	3955.3E6	3647.2E6	753.399	770.976
11) T	2,4,5-TP ...	14.946	14.251	1618.7E6	1478.2E6	745.640	759.581
12) T	2,4,5-T	15.343	14.826	1597.7E6	1117.3E6	742.002	660.693
13) T	2,4-DB	16.096	15.470	211.4E6	183.5E6	717.900	741.422
14) T	DINOSEB	17.607	15.903	1062.9E6	943.9E6	713.800	723.571
15) T	Picloram	17.364	17.147	1829.2E6	1591.7E6	626.346	638.121
16) T	DCPA	17.946	17.089	1925.8E6	1778.0E6	751.550	771.066



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS062818\
Data File : PS000698.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Jun 2018 11:14
Operator : UA
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 29 00:55:06 2018
Quant Method : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061718.M
Quant Title : 8080.M
QLast Update : Mon Jun 18 00:27:44 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 3.174 min
Delta R.T.: -0.008 min
Response: 349459531
Conc: 683.26 ng/ml

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750




#1 Dalapon

R.T.: 2.554 min
Delta R.T.: -0.006 min
Response: 301270022
Conc: 691.09 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 10.228 min
Delta R.T.: -0.015 min
Response: 386415928
Conc: 716.01 ng/ml



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 9.340 min
Delta R.T.: -0.008 min
Response: 331496481
Conc: 720.58 ng/ml



#3 4-Nitrophenol


R.T.: 11.412 min
Delta R.T.: -0.015 min
Response: 161306408
Conc: 686.20 ng/ml

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750




#3 4-Nitrophenol

R.T.: 10.342 min
Delta R.T.: -0.008 min
Response: 135487804
Conc: 691.40 ng/ml




#4 2,4-DCAA

R.T.: 11.814 min
Delta R.T.: -0.015 min
Response: 285388024
Conc: 757.69 ng/ml



#4 2,4-DCAA


R.T.: 11.195 min
Delta R.T.: -0.008 min
Response: 261606118
Conc: 768.73 ng/ml



#5 DICAMBA


R.T.: 12.133 min
Delta R.T.: -0.015 min
Response: 1047089225
Conc: 740.80 ng/ml

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750




#5 DICAMBA

R.T.: 11.484 min
Delta R.T.: -0.009 min
Response: 931052297
Conc: 751.48 ng/ml




#6 MCP

R.T.: 12.472 min
Delta R.T.: -0.010 min
Response: 81736569
Conc: 72.18 ug/ml



#6 MCP


R.T.: 11.709 min
Delta R.T.: -0.002 min
Response: 68568702
Conc: 71.47 ug/ml



#7 MCPA


R.T.: 12.724 min
Delta R.T.: -0.010 min
Response: 122937507
Conc: 69.94 ug/ml

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750




#7 MCPA

R.T.: 12.088 min
Delta R.T.: 0.000 min
Response: 103528430
Conc: 74.75 ug/ml




#8 DICHLORPROP

R.T.: 13.325 min
Delta R.T.: -0.016 min
Response: 269659517
Conc: 716.88 ng/ml



#8 DICHLORPROP


R.T.: 12.608 min
Delta R.T.: -0.008 min
Response: 234128824
Conc: 727.17 ng/ml



#9 2,4-D


R.T.: 13.683 min
Delta R.T.: -0.016 min
Response: 349405753
Conc: 730.51 ng/ml

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750




#9 2,4-D

R.T.: 13.079 min
Delta R.T.: -0.008 min
Response: 316177460
Conc: 743.95 ng/ml




#10 Pentachlorophenol

R.T.: 14.081 min
Delta R.T.: -0.016 min
Response: 3955317120
Conc: 753.40 ng/ml



#10 Pentachlorophenol


R.T.: 13.642 min
Delta R.T.: -0.009 min
Response: 3647234106
Conc: 770.98 ng/ml



#11 2,4,5-TP (SILVEX)


R.T.: 14.946 min
Delta R.T.: -0.016 min
Response: 1618664344
Conc: 745.64 ng/ml

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750




#11 2,4,5-TP (SILVEX)

R.T.: 14.251 min
Delta R.T.: -0.009 min
Response: 1478154403
Conc: 759.58 ng/ml




#12 2,4,5-T

R.T.: 15.343 min
Delta R.T.: -0.017 min
Response: 1597735885
Conc: 742.00 ng/ml



#12 2,4,5-T


R.T.: 14.826 min
Delta R.T.: -0.008 min
Response: 1117319412
Conc: 660.69 ng/ml



#13 2,4-DB


R.T.: 16.096 min
Delta R.T.: -0.018 min
Response: 211362918
Conc: 717.90 ng/ml

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750




#13 2,4-DB

R.T.: 15.470 min
Delta R.T.: -0.008 min
Response: 183537072
Conc: 741.42 ng/ml




#14 DINOSEB

R.T.: 17.607 min
Delta R.T.: -0.018 min
Response: 1062935738
Conc: 713.80 ng/ml



#14 DINOSEB


R.T.: 15.903 min
Delta R.T.: -0.008 min
Response: 943948376
Conc: 723.57 ng/ml



#15 Picloram


R.T.: 17.364 min
Delta R.T.: -0.018 min
Response: 1829229861
Conc: 626.35 ng/ml

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750




#15 Picloram

R.T.: 17.147 min
Delta R.T.: -0.009 min
Response: 1591701259
Conc: 638.12 ng/ml



#16 DCPA

R.T.: 17.946 min
Delta R.T.: -0.018 min
Response: 1925825792
Conc: 751.55 ng/ml



#16 DCPA

R.T.: 17.089 min
Delta R.T.: -0.009 min
Response: 1778048541
Conc: 771.07 ng/ml