

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS063021\
 Data File : PS015988.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Jun 2021 00:46
 Operator : DD\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 30 05:23:12 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS063021.M
 Quant Title : 8080.M
 QLast Update : Wed Jun 30 00:31:48 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.188	6.766	648.9E6	246.1E6	422.700	479.345
Target Compounds							
15) T	Picloram	10.981	10.821	38580638	13807726	3.507	3.319
16) T	DCPA	0.000	10.821f	0	13807726	N.D.	3.211



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS063021\
Data File : PS015988.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Jun 2021 00:46
Operator : DD\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 30 05:23:12 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS063021.M
Quant Title : 8080.M
QLast Update : Wed Jun 30 00:31:48 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#4 2,4-DCAA


R.T.: 7.188 min
Delta R.T.: 0.000 min
Response: 648929527
Conc: 422.70 ng/ml

Instrument :
ECD_S
ClientSampleId :
I.BLK




#4 2,4-DCAA

R.T.: 6.766 min
Delta R.T.: 0.000 min
Response: 246056070
Conc: 479.35 ng/ml




#15 Picloram

R.T.: 10.981 min
Delta R.T.: 0.011 min
Response: 38580638
Conc: 3.51 ng/ml



#15 Picloram


R.T.: 10.821 min
Delta R.T.: 0.005 min
Response: 13807726
Conc: 3.32 ng/ml



#16 DCPA

R.T.: 0.000 min
Exp R.T.: 11.420 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :
I.BLK



#16 DCPA

R.T.: 10.821 min
Delta R.T.: 0.051 min
Response: 13807726
Conc: 3.21 ng/ml