

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS070120\
 Data File : PS011054.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Jul 2020 01:42
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Ankita
 7/2/2020 10:08:31 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 02 07:12:50 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS062420.M
 Quant Title : 8080.M
 QLast Update : Thu Jun 25 02:45:12 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.228	6.920	751.5E6	448.2E6	828.753	808.883
Target Compounds							
1) T	Dalapon	3.041	2.626	768.0E6	495.3E6	685.925	656.640
2) T	3,5-DICHL...	6.441	6.070	971.1E6	626.7E6	771.915	752.471
3) T	4-Nitroph...	7.058	6.525	311.9E6	239.0E6	781.385	743.766
5) T	DICAMBA	7.420	7.089	2845.4E6	1733.9E6	780.336	763.407
6) T	MCPD	7.590	7.187	225.1E6	147.9E6	78.950	78.143
7) T	MCPA	7.740	7.403	310.5E6	211.6E6	77.978	76.657
8) T	DICHLORPROP	8.112	7.714	677.3E6	445.6E6	772.314	758.867
9) T	2,4-D	8.341	8.008	691.1E6	551.7E6	781.512	763.399
10) T	Pentachlo...	8.652	8.439	9631.3E6	5137.4E6	774.584	738.803
11) T	2,4,5-TP ...	9.201	8.799	3866.3E6	2327.7E6	778.655	773.593
12) T	2,4,5-T	9.486	9.179	3630.2E6	2309.2E6	779.364	775.439
13) T	2,4-DB	10.040	9.685	458.4E6	319.0E6	771.358	776.166
14) T	DINOSEB	11.201	10.019	2781.6E6	1644.2E6	776.671	789.668
15) T	Picloram	11.021	11.001	3926.9E6	2919.7E6	700.460	696.896
16) T	DCPA	11.466	10.947	4746.7E6	2917.4E6	776.428m	783.322

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS070120\
Data File : PS011054.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Jul 2020 01:42
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Ankita
7/2/2020 10:08:31 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 02 07:12:50 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS062420.M
Quant Title : 8080.M
QLast Update : Thu Jun 25 02:45:12 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

