

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS070219\
 Data File : PS005357.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Jul 2019 21:07
 Operator : SM\AJ
 Sample : K3600-02
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 WC-6

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 03 00:44:27 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061919.M
 Quant Title : 8080.M
 QLast Update : Thu Jun 20 11:04:04 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.586	5.301	377.8E6	244.0E6	257.363	243.383
Target Compounds							
2) T	3,5-DICHL...	0.000	4.619	0	107.9E6	N.D.	81.858
8) T	DICHLORPROP	0.000	5.989	0	51299794	N.D.	53.131
10) T	Pentachlo...	0.000	6.566	0	16373020	N.D.	1.274
11) T	2,4,5-TP ...	0.000	6.953	0	56360202	N.D.	10.584
12) T	2,4,5-T	7.607	0.000	109.2E6	0	13.567	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS070219\
Data File : PS005357.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Jul 2019 21:07
Operator : SM\AJ
Sample : K3600-02
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
WC-6

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 03 00:44:27 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061919.M
Quant Title : 8080.M
QLast Update : Thu Jun 20 11:04:04 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#2 3,5-DICHLOROBENZOIC ACID


R.T.: 0.000 min
Exp R.T.: 4.914 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :
WC-6




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 4.619 min
Delta R.T.: 0.047 min
Response: 107946266
Conc: 81.86 ng/ml




#4 2,4-DCAA

R.T.: 5.586 min
Delta R.T.: 0.022 min
Response: 377781079
Conc: 257.36 ng/ml



#4 2,4-DCAA


R.T.: 5.301 min
Delta R.T.: 0.021 min
Response: 243959838
Conc: 243.38 ng/ml



#8 DICHLORPROP


R.T.: 0.000 min
Exp R.T.: 6.363 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :
WC-6




#8 DICHLORPROP

R.T.: 5.989 min
Delta R.T.: 0.020 min
Response: 51299794
Conc: 53.13 ng/ml




#10 Pentachlorophenol

R.T.: 0.000 min
Exp R.T.: 6.852 min
Response: 0
Conc: N.D.



#10 Pentachlorophenol


R.T.: 6.566 min
Delta R.T.: -0.053 min
Response: 16373020
Conc: 1.27 ng/ml



#11 2,4,5-TP (SILVEX)


R.T.: 0.000 min
Exp R.T.: 7.387 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :
WC-6




#11 2,4,5-TP (SILVEX)

R.T.: 6.953 min
Delta R.T.: -0.028 min
Response: 56360202
Conc: 10.58 ng/ml



#12 2,4,5-T

R.T.: 7.607 min
Delta R.T.: -0.051 min
Response: 109188367
Conc: 13.57 ng/ml



#12 2,4,5-T

R.T.: 0.000 min
Exp R.T.: 7.341 min
Response: 0
Conc: N.D.