

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS070220\
 Data File : PS011064.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Jul 2020 10:08
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Ankita
 7/6/2020 1:42:01 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 02 14:23:03 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS062420.M
 Quant Title : 8080.M
 QLast Update : Thu Jun 25 02:45:12 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.229	6.921	729.5E6	436.1E6	804.506	787.113
Target Compounds							
1) T	Dalapon	3.041	2.627	763.4E6	489.4E6	681.768	648.775
2) T	3,5-DICHL...	6.442	6.070	936.9E6	608.9E6	744.748	731.087
3) T	4-Nitroph...	7.059	6.526	288.2E6	222.1E6	721.919	691.283
5) T	DICAMBA	7.421	7.089	2761.5E6	1688.4E6	757.330	743.372
6) T	MCPD	7.591	7.188	216.4E6	141.0E6	75.905	74.486
7) T	MCPA	7.740	7.404	299.3E6	204.2E6	75.161	73.986
8) T	DICHLORPROP	8.113	7.715	655.3E6	434.2E6	747.236	739.571
9) T	2,4-D	8.342	8.009	668.1E6	535.7E6	755.484	741.335
10) T	Pentachlo...	8.653	8.440	9354.2E6	5040.1E6	752.298	724.812
11) T	2,4,5-TP ...	9.203	8.800	3753.9E6	2275.4E6	756.033	756.239
12) T	2,4,5-T	9.488	9.179	3521.2E6	2244.9E6	755.956	753.861
13) T	2,4-DB	10.040	9.686	449.4E6	307.8E6	756.073	748.869
14) T	DINOSEB	11.201	10.019	2583.6E6	1518.5E6	721.388	729.278
15) T	Picloram	11.023	11.002	3905.6E6	2902.8E6	696.666	692.877
16) T	DCPA	11.466	10.948	4592.7E6	2821.3E6	751.240m	757.530

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS070220\
 Data File : PS011064.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Jul 2020 10:08
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Ankita
 7/6/2020 1:42:01 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 02 14:23:03 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS062420.M
 Quant Title : 8080.M
 QLast Update : Thu Jun 25 02:45:12 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

