

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS070221\
 Data File : PS016025.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Jul 2021 11:10
 Operator : DD\AJ
 Sample : M2866-01MSD
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 124055MSD

Manual Integrations
 APPROVED

mohammad
 7/2/2021 11:05:24 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 01 16:13:29 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS063021.M
 Quant Title : 8080.M
 QLast Update : Wed Jun 30 00:31:48 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.187	6.766	851.6E6	236.4E6	554.706	460.505
Target Compounds							
1) T	Dalapon	3.014	2.518	779.7E6	254.6E6	359.493	364.273
2) T	3,5-DICHL...	6.405	5.957	515.9E6	164.2E6	237.843	213.311m
3) T	4-Nitroph...	7.014	6.387	160.4E6	60944178	181.407	193.090
5) T	DICAMBA	7.378	6.929	1561.3E6	559.2E6	272.515	265.474
6) T	MCPD	7.544	7.021	94698276	31307194	21.601	18.870
7) T	MCPA	7.692	7.231	155.4E6	74320188	24.431	28.196
8) T	DICHLORPROP	8.071	7.549	309.2E6	113.2E6	222.305	214.937
9) T	2,4-D	8.299	7.839	463.4E6	184.6E6	289.095	292.720
10) T	Pentachlo...	8.608	8.267	4237.9E6	1621.8E6	213.852	225.463
11) T	2,4,5-TP ...	9.159	8.628	1435.5E6	586.5E6	193.451	207.852
12) T	2,4,5-T	9.443	9.006	1606.3E6	616.8E6	205.701	213.822
13) T	2,4-DB	9.992	9.511	267.3E6	84682340	160.935	225.256 #
14) T	DINOSEB	11.152	9.839	760.1E6	289.0E6	138.654	141.526
15) T	Picloram	10.969	10.816	4130.5E6	1621.5E6	375.496	389.741
16) T	DCPA	11.420	10.769	3661.9E6	1671.1E6	339.484	388.594

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS070221\
Data File : PS016025.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Jul 2021 11:10
Operator : DD\AJ
Sample : M2866-01MSD
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
124055MSD

Manual Integrations
APPROVED

mohammad
7/2/2021 11:05:24 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 01 16:13:29 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS063021.M
Quant Title : 8080.M
QLast Update : Wed Jun 30 00:31:48 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

