

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS070818\
 Data File : PS000820.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Jul 2018 16:16
 Operator : UA
 Sample : HSTDICC200
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC200

Manual Integrations
 APPROVED

mohammad
 7/9/2018 2:34:12 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 09 01:56:19 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS070818.M
 Quant Title : 8080.M
 QLast Update : Mon Jul 09 01:45:11 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.635	11.056	44801788	46388084	213.925	208.974
Target Compounds							
1) T	Dalapon	3.026	2.437	50030514	49560586	179.736	178.863
2) T	3,5-DICHL...	10.051	9.202	60281037	59950720	197.802	200.754
3) T	4-Nitroph...	11.228	10.202	26277802	23519211	195.445	191.581
5) T	DICAMBA	11.951	11.343	151.6E6	151.2E6	187.922	186.782
6) T	MCPD	12.286	11.563	8938742	8304701	14.641m	14.507m
7) T	MCPA	12.537	11.940	16493751	13188411	17.479	15.486
8) T	DICHLORPROP	13.143	12.468	40740717	41855058	197.475	204.722
9) T	2,4-D	13.501	12.940	54783777	56124195	198.574	198.956
10) T	Pentachlo...	13.890	13.496	610.0E6	627.9E6	194.555	192.921
11) T	2,4,5-TP ...	14.761	14.110	241.3E6	248.3E6	192.208	191.285
12) T	2,4,5-T	15.157	14.681	241.0E6	203.2E6	192.610	192.463
13) T	2,4-DB	15.909	15.326	31421079	30438438	187.691	192.704
14) T	DINOSEB	17.411	15.758	174.3E6	167.6E6	198.060	195.150
15) T	Picloram	17.168	16.941	357.8E6	305.5E6	192.111	194.460
16) T	DCPA	17.749	16.998	294.7E6	335.4E6	197.614	186.932

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS070818\
Data File : PS000820.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Jul 2018 16:16
Operator : UA
Sample : HSTDICC200
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
HSTDICC200

Manual Integrations
APPROVED

mohammad
7/9/2018 2:34:12 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 09 01:56:19 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS070818.M
Quant Title : 8080.M
QLast Update : Mon Jul 09 01:45:11 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

