

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\
 Method File : PS070818.M
 Title : 8080.M
 Last Update : Mon Jul 09 02:00:25 2018
 Response Via : Initial Calibration

Calibration Files

750 =PS000822.D 1000 =PS000823.D 1500 =PS000824.D
 500 =PS000821.D 200 =PS000820.D

	Compound	750	1000	1500	500	200	Avg		%RSD

1) T	Dalapon	2.784	2.579	2.575	2.739	2.749	2.685	E5	3.72
2) T	3,5-DICHLOROBENZOI...	3.048	2.856	2.755	3.076	3.241	2.995	E5	6.39
3) T	4-Nitrophenol	1.345	1.265	1.319	1.486	1.444	1.372	E5	6.63
4) S	2,4-DCAA	2.094	1.975	1.898	2.109	2.240	2.063	E5	6.40
5) T	DICAMBA	8.069	7.775	7.468	7.991	8.065	7.874	E5	3.25
6) T	MCPP	6.105	6.230	5.592	4.784	4.755	5.493	E5	12.79
7) T	MCPA	9.436	9.321	8.975	9.032	8.868	9.126	E5	2.64
8) T	DICHLORPROP	2.063	1.977	1.816	2.055	2.167	2.016	E5	6.47
9) T	2,4-D	2.759	2.651	2.521	2.765	2.914	2.722	E5	5.37
10) T	Pentachlorophenol	3.136	2.951	2.734	3.155	3.211	3.037	E6	6.44
11) T	2,4,5-TP (SILVEX)	1.256	1.205	1.126	1.252	1.270	1.222	E6	4.81
12) T	2,4,5-T	1.251	1.198	1.126	1.245	1.268	1.218	E6	4.71
13) T	2,4-DB	1.674	1.588	1.531	1.647	1.654	1.619	E5	3.63
14) T	DINOSEB	8.800	8.386	7.940	8.846	9.271	8.648	E5	5.84
15) T	Picloram	1.863	1.770	1.692	1.847	1.883	1.811	E6	4.37
16) T	DCPA	1.491	1.421	1.329	1.504	1.535	1.456	E6	5.66

Signal #2 Calibration Files

750 =PS000822.D 1000 =PS000823.D 1500 =PS000824.D
 500 =PS000821.D 200 =PS000820.D

	Compound	750	1000	1500	500	200	Avg		%RSD

1) T	Dalapon	2.771	2.690	2.608	2.680	2.723	2.694	E5	2.22
2) T	3,5-DICHLOROBENZOI...	2.986	2.809	2.702	2.988	3.223	2.942	E5	6.77
3) T	4-Nitrophenol	1.228	1.160	1.124	1.230	1.292	1.207	E5	5.47
4) S	2,4-DCAA	2.220	2.094	2.030	2.226	2.319	2.178	E5	5.28
5) T	DICAMBA	8.094	7.707	7.500	7.903	8.042	7.849	E5	3.13
6) T	MCPP	5.725	5.561	5.147	5.295	4.417	5.229	E5	9.68
7) T	MCPA	8.516	7.856	7.973	7.429	7.091	7.773	E5	7.00
8) T	DICHLORPROP	2.044	1.927	1.861	2.040	2.226	2.020	E5	6.89
9) T	2,4-D	2.821	2.663	2.584	2.795	2.985	2.770	E5	5.58
10) T	Pentachlorophenol	3.255	3.044	2.854	3.235	3.305	3.139	E6	5.98
11) T	2,4,5-TP (SILVEX)	1.298	1.220	1.172	1.278	1.307	1.255	E6	4.56
12) T	2,4,5-T	1.056	0.965	0.970	1.114	1.069	1.035	E6	6.30
13) T	2,4-DB	1.580	1.475	1.463	1.539	1.602	1.532	E5	4.02
14) T	DINOSEB	8.587	8.082	7.789	8.503	8.914	8.375	E5	5.28
15) T	Picloram	1.571	1.472	1.400	1.563	1.608	1.523	E6	5.57
16) T	DCPA	1.794	1.682	1.627	1.746	1.747	1.719	E6	3.78

(#) = Out of Range