

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS070822\
 Data File : PS019917.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Jul 2022 19:01
 Operator : AJ\MA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 09 01:04:43 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS062722.M
 Quant Title : 8080.M
 QLast Update : Mon Jun 27 17:26:48 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.192	6.777	1923.5E6	582.8E6	741.612	717.548
Target Compounds							
1) T	Dalapon	3.014	2.525	3123.9E6	644.4E6	644.470	607.189
2) T	3,5-DICHL...	6.407	5.959	2373.7E6	855.1E6	692.981	669.161
3) T	4-Nitroph...	7.025	6.397	777.9E6	292.7E6	730.893	676.828
5) T	DICAMBA	7.381	6.941	7183.0E6	2469.8E6	700.024	677.031
6) T	MCPD	7.549	7.038	549.4E6	191.9E6	68.484	66.859
7) T	MCPA	7.698	7.250	810.8E6	267.4E6	69.196	66.072
8) T	DICHLORPROP	8.071	7.560	1882.3E6	632.8E6	695.202	672.698
9) T	2,4-D	8.302	7.853	2075.2E6	751.3E6	698.687	666.814
10) T	Pentachlo...	8.606	8.277	25754.8E6	8477.9E6	713.173	686.629
11) T	2,4,5-TP ...	9.157	8.637	11184.2E6	3647.5E6	699.201	688.258
12) T	2,4,5-T	9.442	9.015	11694.0E6	3397.8E6	635.542	695.517
13) T	2,4-DB	9.995	9.521	2618.0E6	440.2E6	654.240	677.795
14) T	DINOSEB	11.145	9.844	10923.9E6	2543.6E6	671.921	713.642
15) T	Picloram	10.972	10.827	20486.1E6	4227.2E6	656.968	649.984
16) T	DCPA	11.413	10.772	17341.5E6	4491.1E6	672.224	692.983

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS070822\
Data File : PS019917.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Jul 2022 19:01
Operator : AJ\MA
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 09 01:04:43 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS062722.M
Quant Title : 8080.M
QLast Update : Mon Jun 27 17:26:48 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

