

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS070920\
 Data File : PS011136.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Jul 2020 19:16
 Operator : DD\AJ
 Sample : L3203-17MS
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 RO-16MS

Manual Integrations
 APPROVED

Ankita
 7/10/2020 11:24:46 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 10 03:49:42 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS062420.M
 Quant Title : 8080.M
 QLast Update : Thu Jun 25 02:45:12 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.227	6.919	227.3E6	134.3E6	250.667m	242.383m
Target Compounds							
1) T	Dalapon	3.040	2.622	183.4E6	265.0E6	163.749	351.291m#
2) T	3,5-DICHL...	6.440	6.070	198.1E6	123.1E6	157.482	147.752m
3) T	4-Nitroph...	7.058	6.526	98227747	78257613	246.075	243.540
5) T	DICAMBA	7.419	7.088	660.7E6	401.0E6	181.200	176.537
6) T	MCPD	7.583	7.179	37910590	22686003	13.295	11.985
7) T	MCPA	7.732	7.391	65194482	57258154	16.370	20.746m#
8) T	DICHLORPROP	8.111	7.714	132.1E6	96522104	150.634	164.395
9) T	2,4-D	8.340	8.007	176.8E6	147.9E6	199.987	204.612m
10) T	Pentachlo...	8.651	8.438	2239.1E6	1405.1E6	180.077	202.061
11) T	2,4,5-TP ...	9.200	8.797	714.4E6	479.0E6	143.883	159.189m
12) T	2,4,5-T	9.486	9.177	695.2E6	461.8E6	149.240	155.068
13) T	2,4-DB	10.039	9.684	72500957	74353652	121.987	180.915 #
14) T	DINOSEB	11.199	10.020	113.0E6	83419935	31.545m	40.065 #
15) T	Picloram	11.019	10.999	1589.9E6	1042.6E6	283.600	248.854
16) T	DCPA	11.466	10.946	1610.7E6	1163.4E6	263.469	312.377

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS070920\
 Data File : PS011136.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Jul 2020 19:16
 Operator : DD\AJ
 Sample : L3203-17MS
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampled :
 RO-16MS

Manual Integrations
 APPROVED

Ankita
 7/10/2020 11:24:46 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 10 03:49:42 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS062420.M
 Quant Title : 8080.M
 QLast Update : Thu Jun 25 02:45:12 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

