

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071023\
 Data File : PS023662.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Jul 2023 11:10
 Operator : AR\AJ
 Sample : 03507-7MSD
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 03507-7MSD

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 07/11/2023
 Supervised By : Ankita Jodhani 07/11/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 11 05:05:28 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS070723.M
 Quant Title : 8080.M
 QLast Update : Fri Jul 07 18:46:40 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.374	6.811	393.1E6	119.6E6	311.263	285.224
Target Compounds							
1) T	Dalapon	2.478	2.553	379.1E6	199.3E6	269.614	284.717
2) T	3,5-DICHL...	5.746	5.988	373.7E6	114.6E6	192.842	176.908
3) T	4-Nitroph...	6.210	6.421	284.5E6	81952620	317.943	284.141m
5) T	DICAMBA	6.517	6.975	1194.4E6	314.8E6	214.130	173.771
8) T	DICHLORPROP	7.107	7.594	278.8E6	84678431	188.510	181.980
9) T	2,4-D	7.305	7.879	381.6E6	125.5E6	220.773m	232.037
10) T	Pentachlo...	7.536	8.307	3095.8E6	904.4E6	147.516	135.067
11) T	2,4,5-TP ...	8.061	8.666	1213.7E6	329.6E6	145.221	131.978m
12) T	2,4,5-T	8.316	9.038	1447.5E6	387.2E6	179.081	158.678
13) T	2,4-DB	8.825	9.543	239.0E6	54117741	242.719	173.929 #
14) T	DINOSEB	9.881	9.872	679.2E6	202.6E6	116.424	114.949
15) T	Picloram	9.704	10.831	3361.0E6	1117.6E6	286.645	264.201
16) T	DCPA	10.145	10.796	4073.6E6	1180.8E6	433.003	425.625

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071023\
Data File : PS023662.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Jul 2023 11:10
Operator : AR\AJ
Sample : 03507-7MSD
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
03507-7MSD

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 07/11/2023
Supervised By : Ankita Jodhani 07/11/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 11 05:05:28 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS070723.M
Quant Title : 8080.M
QLast Update : Fri Jul 07 18:46:40 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

