

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071118\
 Data File : PS000939.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Jul 2018 13:53
 Operator : UA
 Sample : J3879-07MS
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 SURCHARGE-SP-4MS

Manual Integrations
 APPROVED

Sohil
 7/12/2018 4:00:38 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 11 15:14:48 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS070818.M
 Quant Title : 8080.M
 QLast Update : Mon Jul 09 02:00:25 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.679	11.064	1505400	1835346	7.296m	8.427m
Target Compounds							
1) T	Dalapon	3.059	2.467	25447275	20430312	94.773	75.829m
2) T	3,5-DICHL...	10.090	9.234	27646345	27353398	92.302m	92.990
3) T	4-Nitroph...	11.267	10.234	18475501	17347861	134.701m	143.769m
5) T	DICAMBA	11.991	11.375	74694053	70388745	94.865	89.674m
6) T	MCP	12.329	11.551	1701510	9287644	3.098	17.762 #
7) T	MCPA	12.577	11.972	5691425	2849747	6.236	3.666 #
8) T	DICHLORPROP	13.183	12.501	17907761	22392947	88.843m	110.868
9) T	2,4-D	13.541	12.972	29071491	28883327	106.806	104.287
10) T	Pentachlo...	13.931	13.528	258.2E6	242.6E6	85.006	77.289
11) T	2,4,5-TP ...	14.800	14.141	89500035	82386790	73.257	65.651m
12) T	2,4,5-T	15.199	14.712	83764350	81499355	68.788m	78.766m
13) T	2,4-DB	15.951	15.358	12339083	12416572	76.222m	81.059m
14) T	DINOSEB	17.451	15.788	17413638	17538058	20.135	20.941m
15) T	Picloram	17.210	16.974	194.6E6	137.2E6	107.441	90.136
16) T	DCPA	17.794	17.031	108.4E6	193.4E6	74.446	112.512 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071118\
Data File : PS000939.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Jul 2018 13:53
Operator : UA
Sample : J3879-07MS
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
SURCHARGE-SP-4MS

Manual Integrations
APPROVED

Sohil
7/12/2018 4:00:38 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 11 15:14:48 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS070818.M
Quant Title : 8080.M
QLast Update : Mon Jul 09 02:00:25 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

